



Agilent 1290 Infinity II Diode Array Detector

User Manual



Agilent Technologies

Notices

© Agilent Technologies, Inc. 2014

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

G7117-90000 Rev. B

Edition

12/2014

Printed in Germany

Agilent Technologies
Hewlett-Packard-Strasse 8
76337 Waldbronn

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as “Commercial computer software” as defined in DFAR 252.227-7014 (June 1995), or as a “commercial item” as defined in FAR 2.101(a) or as “Restricted computer software” as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies’ standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will

receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

In This Book

This manual covers the Agilent 1290 Infinity II Diode Array Detectors

- G7117B DAD (Variable Slit)
- G7117A DAD-FS (Fixed Slit)

Find information on other Agilent Diode Array Detectors in separate manuals.

1 Introduction

This chapter gives an introduction to the detector and an instrument overview.

2 Site Requirements and Specifications

This chapter provides information on environmental requirements, physical and performance specifications.

3 Using the Module

This chapter explains the essential operational parameters of the module.

4 Preparing the Module

This chapter provides information on how to set up the module for an analysis and explains the basic settings.

5 Optimizing the Detector

This chapter provides information on how to optimize the detector.

6 Troubleshooting and Diagnostics

Overview about the troubleshooting and diagnostic features.

7 Error Information

This chapter describes the meaning of error messages, and provides information on probable causes and suggested actions how to recover from error conditions.

8 Test Functions and Calibration

This chapter describes the tests for the module.

9 Maintenance

This chapter describes the maintenance of the module.

10 Parts and Materials for Maintenance

This chapter provides information on parts for maintenance.

11 Identifying Cables

This chapter provides information on cables used with the Agilent 1200 Infinity Series modules.

12 Hardware Information

This chapter describes the detector in more detail on hardware and electronics.

13 LAN Configuration

This chapter provides information on connecting the module to the Agilent ChemStation PC.

14 Appendix

This chapter provides addition information on safety, legal and web.

Contents

1	Introduction	9
	Overview of the Module	10
	Product Description	11
	Optical System	13
	Leak and Waste Handling	18
	Operating Principle	21
2	Site Requirements and Specifications	23
	Site Requirements	24
	Physical Specifications	27
	Performance Specifications	28
3	Using the Module	33
	Magnets	34
	Turn on/off	35
	Status Indicators	36
	Instrument Configuration	37
	Set up the Detector with Agilent Open Lab ChemStation	39
	The Detector User Interface	40
	Detector Control Settings	42
	Method Parameter Settings	43
4	Preparing the Module	49
	Leak and Waste Handling	50
	Setting up an Analysis	52
	Solvent Information	59

5	Optimizing the Detector	65
	Introduction	66
	Optimization Overview	67
	Optimizing for Sensitivity, Selectivity, Linearity and Dispersion	72
	Optimizing Selectivity	83
	Optimizing the Detector Regarding to the System	87
	Warm up of the Detector	94
6	Troubleshooting and Diagnostics	97
	Available Tests vs User Interfaces	98
	Agilent Lab Advisor Software	99
7	Error Information	101
	What Are Error Messages	103
	General Error Messages	104
	Detector Error Messages	111
8	Test Functions and Calibration	117
	Introduction	118
	Use of Max-Light Cartridge Test Cell	120
	Conditions of Detector	121
	Failing a Test	121
	Self-Test	122
	Intensity Test	124
	Cell Test	127
	Quick Noise Test	130
	ASTM Drift and Noise Test	133
	Slit Test	136
	Wavelength Verification Test	139
	Wavelength Calibration	141
	D/A Converter (DAC) Test	144
	Dark Current Test	147
	Other Lab Advisor Functions	149
9	Maintenance	151
	Introduction to Maintenance	152
	Warnings and Cautions	152


Overview of Maintenance	154
Cleaning the Module	155
Remove and Install Doors	156
Replace the Deuterium Lamp	157
Replace the Max-Light Cartridge Cell	160
Clean the Max-Light Cartridge Cell	163
Storage of Max-Light Cartridge Cell	165
Correcting Leaks	166
Replace Leak Handling System Parts	167
Replace the Module Firmware	169
Information from Module's Assemblies	171
10 Parts and Materials for Maintenance	173
Overview of Maintenance Parts	174
Accessory Kit	176
11 Identifying Cables	177
Cable Overview	178
Analog Cables	180
Remote Cables	182
CAN/LAN Cables	186
RS-232 Cables	187
USB Cables	187
12 Hardware Information	189
Firmware Description	190
Electrical Connections	193
Interfaces	196
Setting the 6-bit Configuration Switch	203
Instrument Layout	207
Early Maintenance Feedback	208

13 LAN Configuration 211

- What You Have to Do First 212
- TCP/IP parameter configuration 213
- Configuration Switches 214
- Initialization Mode Selection 215
- Dynamic Host Configuration Protocol (DHCP) 219
- Manual Configuration 222
- PC and Agilent ChemStation Setup 226

14 Appendix 235

- General Safety Information 236
- Waste Electrical and Electronic Equipment Directive 244
- Radio Interference 245
- Sound Emission 245
- Solvent Information 246
- Agilent Technologies on Internet 247



1

Introduction

Overview of the Module	10
Product Description	11
Product Description G7117A	11
Product Description G7117B	12
Optical System	13
Lamp	14
Max-Light Cartridge Flow Cell	14
Slit Assembly	15
Grating and Diode Array	17
Leak and Waste Handling	18
Waste Concept	20
Operating Principle	21

This chapter gives an introduction to the detector and an instrument overview.



Overview of the Module

The detector is designed for highest optical performance, GLP compliance and easy maintenance. It includes the following features:

- Maximum of 120 Hz (G7117A) or 240 Hz (G7117B) data acquisition rate.
- Higher sensitivity for conventional LC as well as ultra fast applications by using next generation optical design.
- Increased sensitivity with 60 mm Max-Light cartridge flow cell.
- Optimized cell geometry for less peak dispersion for narrow bore applications.
- Max-Light cartridge flow cells for standard applications are available, see [“Max-Light Cartridge Flow Cell”](#) on page 14.
- More reliable and robust peak integration process (automated) due to less baseline noise/drift/refractive index and thermal effects especially under ultra fast gradient conditions.
- RFID tracking technology is used for the UV-lamp and the Max-Light cartridge flow cells.
- Multiple wavelength and full spectral detection at 120 Hz (G7117A)/240 Hz (G7117B) sampling rate, keeping up with the analysis speed of ultra-fast LC.
- Programmable 1 – 8 nm slit (G7117B) or fixed 4 nm slit (G7117B) for rapid optimization of sensitivity, linearity and spectral resolution provides optimum incident light conditions.
- Improved Electronic temperature control (ETC) provides maximum baseline stability and practical sensitivity under fluctuating ambient temperature and humidity conditions.
- Additional diagnostic signals for temperature and lamp voltage monitoring.
- Easy exchange of flow cell by cartridge design.

Product Description

Product Description G7117A

The Agilent 1290 Infinity II Diode Array Detector FS (fixed slit) is based on the Agilent Max-Light cartridge cell with optofluidic waveguides that improve light transmission to near 100% efficiency without sacrificing resolution caused by cell dispersions effects.

With typical detector noise levels of $< \pm 0.6 \mu\text{AU}/\text{cm}$ the 60 mm flow cell gives up to 10 times higher sensitivity than detectors with conventional flow cells.

Any compromising refractive index and thermal effects are almost completely eliminated, resulting in significantly less baseline drift for more reliable and precise peak integration.

For fast separations, this detector has multiple wavelength and full spectral detection at sampling rates up to 120 Hz.




Figure 1 Overview of the Diode Array Detector

Product Description G7117B

The Agilent 1290 Infinity II Diode Array Detector (DAD) is based on the Agilent Max-Light cartridge cell with optofluidic waveguides that improve light transmission to near 100% efficiency without sacrificing resolution caused by cell dispersions effects.

With typical detector noise levels of $< \pm 0.6 \mu\text{AU}/\text{cm}$ the 60 mm flow cell gives up to 10 times higher sensitivity than detectors with conventional flow cells.

Any compromising refractive index and thermal effects are almost completely eliminated, resulting in significantly less baseline drift for more reliable and precise peak integration.

For fast separations, this detector has multiple wavelength and full spectral detection at sampling rates up to 240 Hz.




Figure 2 Overview of the Diode Array Detector

Optical System

The optical system of the detector is shown in [Figure 3](#) on page 13.




Figure 3 Optical System of the Detector

1	UV-lamp
2	Lamp mirror
3	Flow cell
4	Fold mirror
5	Programmable (G7117B) or Fixed (G7117A) slit
6	Grating
7	Array

The illumination source is a deuterium-arc-discharge lamp [1] for the ultraviolet (UV) wavelength range. Its light is focused by a lamp mirror [2] onto the entrance of the Max-light cartridge flow cell [3] with optofluidic waveguides. The light leaves the Max-light cartridge flow cell at the other side and is focused by the fold mirror [4] through the slit assembly [5] onto a holographic grating [6] light being dispersed onto the diode array [7]. This allows simultaneous access to all wavelength information.

Lamp

The light source for the UV-wavelength range is a long-life UV-lamp with RFID tag. As a result of plasma discharge in low-pressure deuterium gas, the lamp emits light over the 190 nm to approximately 800 nm wavelength range.




Figure 4 UV-Lamp

Max-Light Cartridge Flow Cell

The detector allows easy access to flow cells via a cartridge. A variety of optional flow cells can be inserted using the same quick, simple mounting system.

Max-Light Cartridge Flow Cells for standard and bio-inert applications are available. For testing of the detector, a Max-Light Cartridge Test Cell is available.

p/n	Description
G4212-60008	Max-Light Cartridge Cell (10 mm, V(σ) 1.0 μ L)
G4212-60007	Max-Light Cartridge Cell (60 mm, V(σ) 4.0 μ L)
G4212-60032	HDR Max-Light Cartridge Cell (3.7 mm, V(σ) 0.4 μ L)
G4212-60038	ULD Max-Light Cartridge Cell (10 mm, V(σ) 0.6 μ L)
G4212-60011	Max-Light Cartridge Test Cell

The optical principle of the Max-Light Cartridge cell is based on opto-fluidic waveguides. Nearly 100 % light transmission is achieved by utilizing total internal reflection in a non-coated silica fiber. Compromising refractive index and thermal effects are almost completely eliminated, resulting in significantly less baseline drift.




Figure 5 Max-Light Cartridge Flow Cell

NOTE

For additional information on the Max-Light Cartridge flow cell refer to [“Choosing a Flow Cell”](#) on page 68 and [“Inline Pressure Relief Valve Kit \(G4212-68001\)”](#) on page 70.

Slit Assembly

Programmable Slit (G7117B)

The micro-slit system makes use of the mechanical properties of silicon combined with the precise structuring capabilities of bulk micro-machining. It combines the required optical functions – slit and shutter – in a simple and compact component. The slit width is directly controlled by the micro-processor of the instrument and can be set as method parameter.



Figure 6 Slit Assembly

The slit width influences the spectral resolution and noise.




Figure 7 Influence of slitwidth on resolution and noise level

Fixed Slit (G7117A)

The fixed slit combines the required optical functions - slit and shutter - in a simple and compact component. The slit width is fixed to 4 nm and both positions (fixed slit and shutter) are directly controlled by the micro-processor of the instrument.

Grating and Diode Array

The combination of dispersion and spectral imaging is accomplished by using a concave holographic grating. The grating separates the light beam into all its component wavelengths and reflects the light onto the photodiode array.

The diode array is a series of 1024 individual photodiodes and control circuits located on a ceramic carrier. It has a wavelength range from 190 – 640 nm and the sampling interval is ~0.5 nm.



Figure 8 Grating and diode array

Leak and Waste Handling

The 1290 Infinity II Series has been designed for safe leak and waste handling. It is important that all security concepts are understood and instructions are carefully followed.




Figure 9 Leak and waste handling concept (overview - typical stack configuration as an example)

The solvent cabinet is designed to store a maximum volume of 6 L solvent. The maximum volume for an individual bottle stored in the solvent cabinet should not exceed 4 L. For details, see the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets (a printed copy of the guideline has been shipped with the solvent cabinet, electronic copies are available on the Internet).

All leak plane outlets are situated in a consistent position so that all Infinity and Infinity II modules can be stacked on top of each other. Waste tubes are guided through a channel on the right hand side of the instrument, keeping the front access clear from tubes.

The leak plane provides leak management by catching all internal liquid leaks, guiding them to the leak sensor for leak detection, and passing them on to the next module below, if the leak sensor fails. The leak sensor in the leak plane stops the running system as soon as the leak detection level is reached.

Solvent and condensate is guided through the waste channel into the waste container:

- from the detector's flow cell outlet
- from the Multisampler needle wash port
- from the Sample Cooler (condensate)
- from the Seal Wash Sensor
- from the pump's Purge Valve or Multipurpose Valve

The waste tube connected to the leak pan outlet on each of the bottom instruments guides the solvent to a suitable waste container.

Waste Concept

- 1** Agilent recommends using the 6 L waste can with 1 Stay Safe cap GL45 with 4 ports (5043-1221) for optimal and safe waste disposal. If you decide to use your own waste solution, make sure that the tubes don't immerse in the liquid.



Operating Principle





Figure 10 Hydraulic path

1 Introduction

Operating Principle



2 Site Requirements and Specifications

Site Requirements	24
Physical Specifications	27
Performance Specifications	28
Specifications	28
Specification Conditions	32

This chapter provides information on environmental requirements, physical and performance specifications.



Site Requirements

A suitable environment is important to ensure optimal performance of the module.

Power Consideration

The module power supply has wide ranging capabilities and accepts any line voltage in the range mentioned in [Table 1](#) on page 27. Consequently, there is no voltage selector in the rear of the module. There are also no externally accessible fuses, because automatic electronic fuses are implemented in the power supply.

WARNING

Module is partially energized when switched off, as long as the power cord is plugged in.

Repair work at the module can lead to personal injuries, e.g. shock hazard, when the cover is opened and the module is connected to power.

- Make sure that it is always possible to access the power plug.
- Remove the power cable from the instrument before opening the cover.
- Do not connect the power cable to the Instrument while the covers are removed.

WARNING

Incorrect line voltage at the module

Shock hazard or damage of your instrument can result if the devices are connected to line voltage higher than specified.

- Connect your module to the specified line voltage.

CAUTION

Inaccessible power plug.

In case of emergency it must be possible to disconnect the instrument from the power line at any time.

- Make sure the power connector of the instrument can be easily reached and unplugged.
- Provide sufficient space behind the power socket of the instrument to unplug the cable.

Power Cords

Different power cords are offered as options with the module. The female end of all power cords is identical. It plugs into the power-input socket at the rear. The male end of each power cord is different and designed to match the wall socket of a particular country or region.

WARNING

Absence of ground connection or use of unspecified power cord

The absence of ground connection or the use of unspecified power cord can lead to electric shock or short circuit.

- Never operate your instrumentation from a power outlet that has no ground connection.
 - Never use a power cord other than the Agilent Technologies power cord designed for your region.
-

WARNING

Use of unsupplied cables

Using cables not supplied by Agilent Technologies can lead to damage of the electronic components or personal injury.

- Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.
-

WARNING

Unintended use of supplied power cords

Using power cords for unintended purposes can lead to personal injury or damage of electronic equipment.

- Never use the power cords that Agilent Technologies supplies with this instrument for any other equipment.
-

Bench Space

The module dimensions and weight (see [Table 1](#) on page 27) allow you to place the module on almost any desk or laboratory bench. It needs an additional 2.5 cm (1.0 inches) of space on either side and approximately 8 cm (3.1 inches) in the rear for air circulation and electric connections.

If the bench shall carry a complete HPLC system, make sure that the bench is designed to bear the weight of all modules.

The module should be operated in a horizontal position.

Environment

Your module will work within the specifications at ambient temperatures and relative humidity described in [Table 1](#) on page 27.

ASTM drift tests require a temperature change below 2 °C/hour (3.6 F/hour) over one hour period. Our published drift specification (refer also to [“Specifications”](#) on page 28) is based on these conditions. Larger ambient temperature changes will result in larger drift.

Better drift performance depends on better control of the temperature fluctuations. To realize the highest performance, minimize the frequency and the amplitude of the temperature changes to below 1 °C/hour (1.8 F/hour). Turbulences around one minute or less can be ignored.

CAUTION

Condensation within the module

Condensation can damage the system electronics.

- Do not store, ship or use your module under conditions where temperature fluctuations could cause condensation within the module.
- If your module was shipped in cold weather, leave it in its box and allow it to warm slowly to room temperature to avoid condensation.

NOTE

This module is designed to operate in a typical electromagnetic environment, i.e. where RF transmitters such as mobile telephones may not be used in close proximity.

Physical Specifications

Table 1 Physical Specifications

Type	Specification	Comments
Weight	11.5 kg (25.4 lbs)	
Dimensions (height × width × depth)	140 x 396 x 436 mm (5.5 x 15.6 x 17.2 inches)	
Line voltage	100 – 240 V~, ± 10 %	Wide-ranging capability
Line frequency	50 or 60 Hz, ± 5 %	
Power consumption	110 VA, 100 W	
Ambient operating temperature	4 – 40 °C (39 – 104 °F)	
Ambient non-operating temperature	-40 – 70 °C (-40 – 158 °F)	
Humidity	< 95 % r.h. at 40 °C (104 °F)	Non-condensing
Operating altitude	Up to 2000 m (6562 ft)	
Non-operating altitude	Up to 4600 m (15092 ft)	For storing the module
Safety standards: IEC, EN, CSA, UL	Installation category II, Pollution degree 2	For indoor use only.

Performance Specifications

Specifications

Performance Specifications G7117B

Table 2 Agilent 1290 Infinity II Diode Array Detector (G7117B) Performance Specifications

Feature	Specification
Detector type	1024-element diode array
Light source	Deuterium
Number of signals	8
Maximum sampling rate	240 Hz (both spectra and signals)
Short-term noise	with 10 mm Max-Light cartridge cell: $<\pm 3 \cdot 10^{-6}$ AU at 230/4 nm, slit width 4 nm, TC 2 s, ASTM with 60 mm Max-Light cartridge cell: $<\pm 0.6 \cdot 10^{-6}$ AU/cm at 230/4 nm, slit width 4 nm, TC 2 s, ASTM
Drift	$<0.5 \cdot 10^{-3}$ AU/h at 230 nm
Linearity	>2.0 AU (5 %) at 265 nm Typically 2.5 AU (5 %)
Wavelength range	190 – 640 nm
Wavelength accuracy	± 1 nm, self-calibration with deuterium lines
Wavelength precision	$<\pm 0.1$ nm
Slit width	Programmable: 1, 2, 4, 8 nm
Diode width	~ 0.5 nm
Wavelength bunching	Programmable, 2 – 400 nm, in steps of 1 nm

Table 2 Agilent 1290 Infinity II Diode Array Detector (G7117B) Performance Specifications

Feature	Specification
Spectral tools	Data analysis software for spectra evaluation, including spectral libraries and peak purity functions
Flow cells	User-exchangeable, self-aligning cartridge cells with RFID tags. Max-Light Cartridge Cell (Standard): 10 mm, $\sigma V = 1.0 \mu L$ Max-Light Cartridge Cell (High Sensitivity): 60 mm, $\sigma V = 4 \mu L$ Max-Light Cartridge Ultra Low Dispersion (ULD) Cell: 10 mm, $\sigma V = 0.6 \mu L$ Max-Light Cartridge High Dynamic Range (HDR) Cell: 3.7 mm, $\sigma V = 0.8 \mu L$ Maximum Operating Pressure (MOP) ¹ : 70 bar Maximum Incidental Pressure (MIP) ² : 150 bar
Analog output	Recorder/integrator: 100 mV or 1 V, output range 0.001 – 2 AU, one output
Communications	LAN, controller-area network (CAN), ERI: ready, start, stop and shut-down signals
GLP features	Data recovery card to prevent data losses. RFID for electronics records of flow cell and UV lamp conditions (path length, volume, product number, serial number, test passed, usage) Early maintenance feedback (EMF) for continuous tracking of instrument usage in terms of lamp burn time with user settable limits and feedback messages. Electronic records of maintenance and errors. Verification of wavelength accuracy with deuterium lines.
Safety and maintenance	Extensive diagnostics, error detection and display through Agilent Instant Pilot and Agilent Lab Advisor software. Leak detection, safe leak handling, leak output signal for shutdown of pumping system. Low voltages in major maintenance areas.
Others	Second generation of Electronic temperature control (ETC) for the complete optical unit

¹ Maximum operating pressure (MOP): Maximum pressure at which a system can operate continuously under normal conditions.

² Maximum incidental pressure (MIP): The maximum pressure which the system can experience during a short time.

Performance Specifications G7117A

Table 3 Agilent 1290 Infinity II Diode Array Detector FS (G7117A) Performance Specifications

Feature	Specification
Detector type	1024-element diode array
Light source	Deuterium
Number of signals	8
Maximum sampling rate	120 Hz (both spectra and signals)
Short-term noise	with 10 mm Max-Light cartridge cell: $<\pm 3 \cdot 10^{-6}$ AU at 230/4 nm, slit width 4 nm, TC 2 s, ASTM with 60 mm Max-Light cartridge cell: $<\pm 0.6 \cdot 10^{-6}$ AU/cm at 230/4 nm, slit width 4 nm, TC 2 s, ASTM
Drift	$<0.5 \cdot 10^{-3}$ AU/h at 230 nm
Linearity	>2.0 AU (5 %) at 265 nm Typically 2.5 AU (5 %)
Wavelength range	190 – 640 nm
Wavelength accuracy	± 1 nm, self-calibration with deuterium lines
Wavelength precision	$<\pm 0.1$ nm
Diode width	~ 0.5 nm
Wavelength bunching	Programmable, 2 – 400 nm, in steps of 1 nm
Flow cells	User-exchangeable, self-aligning cartridge cells with RFID tags. Max-Light Cartridge Cell (Standard): 10 mm, $\sigma V = 1.0$ μ L Max-Light Cartridge Cell (High Sensitivity): 60 mm, $\sigma V = 4$ μ L Max-Light Cartridge Ultra Low Dispersion (ULD) Cell: 10 mm, $\sigma V = 0.6$ μ L Max-Light Cartridge High Dynamic Range (HDR) Cell: 3.7 mm, $\sigma V = 0.8$ μ L Maximum Operating Pressure (MOP) ¹ : 70 bar Maximum Incidental Pressure (MIP) ² : 150 bar

Table 3 Agilent 1290 Infinity II Diode Array Detector FS (G7117A) Performance Specifications

Feature	Specification
Spectral tools	Data analysis software for spectra evaluation, including spectral libraries and peak purity functions
Analog output	Recorder/integrator: 100 mV or 1 V, output range 0.001 – 2 AU, one output
Communications	LAN, controller-area network (CAN), ERI: ready, start, stop and shut-down signals
GLP features	Data recovery card to prevent data losses. RFID for electronics records of flow cell and UV lamp conditions (path length, volume, product number, serial number, test passed, usage) Early maintenance feedback (EMF) for continuous tracking of instrument usage in terms of lamp burn time with user settable limits and feedback messages. Electronic records of maintenance and errors. Verification of wavelength accuracy with deuterium lines.
Safety and maintenance	Extensive diagnostics, error detection and display through Agilent Instant Pilot and Agilent Lab Advisor software. Leak detection, safe leak handling, leak output signal for shutdown of pumping system. Low voltages in major maintenance areas.
Others	Second generation of Electronic temperature control (ETC) for the complete optical unit

¹ Maximum operating pressure (MOP): Maximum pressure at which a system can operate continuously under normal conditions.

² Maximum incidental pressure (MIP): The maximum pressure which the system can experience during a short time.

Specification Conditions

ASTM: “Standard Practice for Variable Wavelength Photometric Detectors Used in Liquid Chromatography”.

Reference conditions:

- Wavelength: 230 nm/4 nm with Reference Wavelength 360 nm/100 nm, Slitwidth 4 nm, TC 2 s, (or with $RT = 2.2 * TC$), ASTM
- Max-Light Cartridge Cell (10 mm, V(s) 1.0 μ L) (G4212-60008) with flow of 0.5 mL/min LC grade water or Max-Light Cartridge Test Cell (G4212-60011)

Linearity:

Linearity is measured with caffeine at 265 nm/4 nm with slit width 4 nm and TC 1 s (or with RT 2 s) with Max-Light Cartridge Cell (10 mm, V(s) 1.0 μ L) (G4212-60008) > 2.0 AU (5 %) [typical 2.5 AU (5 %)] .

NOTE

The specifications are based on the standard RFID tag lamp (5190-0917) and may be not achieved when other lamp types or aged lamps are used.


ASTM drift tests require a temperature change below 2 °C/hour (3.6 F/hour) over one hour period. Our published drift specification is based on these conditions. Larger ambient temperature changes will result in larger drift.

Better drift performance depends on better control of the temperature fluctuations. To realize the highest performance, minimize the frequency and the amplitude of the temperature changes to below 1 °C/hour (1.8 F/hour). Turbulences around one minute or less can be ignored.

Performance tests should be done with a completely warmed up optical unit (> two hours). ASTM measurements require that the detector should be turned on at least 24 h before start of testing.

Time Constant versus Response Time

According to ASTM E1657-98 „Standard Practice of Testing Variable- Wavelength Photometric Detectors Used in Liquid Chromatography” the time constant is converted to response time by multiplying by the factor 2.2.




3 Using the Module

Magnets	34
Turn on/off	35
Status Indicators	36
Instrument Configuration	37
Set up the Detector with Agilent Open Lab ChemStation	39
The Detector User Interface	40
Detector Control Settings	42
Method Parameter Settings	43



This chapter explains the essential operational parameters of the module.



Magnets




Turn on/off

1 	2  Power switch (1) On (2) Off
3 	

Status Indicators

- 1** The module status indicator indicates one of six possible module conditions:




Status indicators


1. Idle
2. Run mode
3. Not-ready. Waiting for a specific pre-run condition to be reached or completed.
4. Error mode - interrupts the analysis and requires attention (for example a leak or defective internal components).
5. Resident mode (blinking) - for example during update of main firmware.
6. Bootloader mode (fast blinking). Try to re-boot the module or try a cold-start. Then try a firmware update.

Instrument Configuration

- 1 Set the switches of the Configuration switch at the rear of the module:
 - a All switches DOWN: module uses the default IP address 192.168.254.11.
 - b Switch 4 UP and others DOWN: module uses DHCP.




- 2 Enter the setup information (MAC / IP address and/or Instrument Name).
 - a Agilent OpenLab ChemStation (Configure Instrument):



3 Using the Module

Instrument Configuration

b Lab Advisor (Instrument Overview - Add Instrument):



The screenshot shows the 'System Properties' dialog box with the 'Instruments' tab selected. The 'System' section at the top has a 'System Name' field containing 'G7117B' and an empty 'Description' field. Below this, the 'Instruments' section contains a table with three columns: 'Instrument Name', 'Instrument Address', and 'Instrument Type'. The first row of the table has 'G7117B' in the 'Instrument Name' column, '0030D3285853' in the 'Instrument Address' column, and 'Agilent LC/CE' in the 'Instrument Type' column. A dropdown menu is open for the 'Instrument Address' column, showing '192.168.254.11' as an option. To the left of the table is a small icon of a laboratory instrument. Below the table is an 'Add Instrument' button. At the bottom of the dialog box, there is a 'Reconnect' checkbox (unchecked), an 'Apply' button, and a 'Cancel' button.

Instrument Name	Instrument Address	Instrument Type
G7117B	0030D3285853	Agilent LC/CE

Set up the Detector with Agilent Open Lab ChemStation

The setup of the detector is shown with the Agilent OpenLab ChemStation C.01.06.

NOTE

This section describes the detector settings only. For information on the Agilent OpenLab ChemStation or other 1200 Infinity modules refer to the corresponding documentation.





Figure 11 ChemStation Method and Run Control (just detector is shown)


After successful load of the OpenLab ChemStation, you should see the module as an active item in the graphical user interface (GUI).

The Detector User Interface



Within the detector GUI, there are active areas. If you move the mouse cursor across the icons the cursor will change.

- 1 Lamp: turn on and off of UV-lamp
- 2 EMF status
- 3 Detector status
- 4 Lamp status (on/off) and information (RFID tag)
- 5 Flow Cell information (RFID tag)



RFID tag information is displayed when moving with the mouse cursor on to the tag attached to the flow cell or lamp. The information provides flow cell and lamp related information like

- Part number
 - Production date
 - Serial number
- and other details.



EMF Status shows Run / Ready / Error state and "Not Ready text" or "Error text"


- Offline (gray)
- Ok. No Maintenance required (green)
- EMF warning. Maintenance might be required (yellow)
- EMF warning. Maintenance required (red)

Important: The EMF settings can be accessed via Agilent Lab Advisor. The limit(s) can be changed. Based on the limit, the User Interface displays the above status.



Module Status shows Run / Ready / Error state and “Not Ready text” or “Error text”


- Error (red)
- Not ready (yellow)
- Ready (green)
- Pre run, Post run (purple)
- Run (blue)
- Idle (green)
- Offline (dark gray)
- Standby (light gray)



A right-click into the Active Area will open a menu to

- Show the Control Interface (special module settings)
- Show the Method interface (similar as via menu **Instrument > Setup Instrument Method**)
- Set Error Method
- Identify Module (Status LED will blink)
- Perform a Balance
- Switch the UV-lamp on/off (same as click on button “Make Device Ready/Turn device off (standby)”)

Detector Control Settings



The figure shows the default settings.

- **Lamps:** can be turned ON/OFF.
- **Analog Output Range:** can be set to either 100 mV or 1 V full scale, for additional settings see **Analog Output** (under “[Method Parameter Settings](#)” on page 43).
- **UV Lamp Tag**
 - **Automatic** detects a lamp with RFID tag. If no RFID tag lamp is used, “UV lamp not ready” is displayed and it cannot be ignited. A compatible mode has to be selected based on the used lamp; see Non-RFID-tag lamp information below.
 - **Manual (by PN)** uses the selected “heating” mode. This mode can also be used when the RFID tag of the standard lamp (Long-life Deuterium lamp (8-pin) with RFID tag (5190-0917)) is not recognized (defect RFID tag or reader).
 - **Non-RFID-tag lamp:** In case a non-RFID-tag lamp is used, the user interface will show this when selecting a compatible mode. You may operate the detector outside of the guaranteed specification. The correct selection is important for optimal performance and lifetime.
- **At Power On:** automatic lamp-on at power on.
- **Automatic Turn On:** automatic detector power on.

Method Parameter Settings

These settings are available via **Menu > Instrument > Set up Instrument Method** or via right click into the module's active area (does not show the **Instrument Curves** tab).

Setup Method

DAD (G7117B)

Signals

	Use Signal	Wave length	Band width	Reference Wavelength	Reference Bandwidth
Signal A	☒	254.0	4.0	☒ 360.0	100.0 nm
Signal B	☒	210.0	4.0	☒ 360.0	100.0 nm
Signal C	☒	214.0	4.0	☒ 360.0	100.0 nm
Signal D	☒	230.0	4.0	☒ 360.0	100.0 nm
Signal E	☒	260.0	4.0	☒ 360.0	100.0 nm
Signal F	☐	273.0	4.0	☒ 360.0	100.0 nm
Signal G	☐	280.0	4.0	☒ 360.0	100.0 nm
Signal H	☐	250.0	4.0	☒ 360.0	100.0 nm

Peakwidth

>0.10 min (2.0 s response time) (2.5 Hz)

Stoptime

☒ As Pump/Injector

☐ 1.00 min

Posttime

☐ Off

☐ 1.00 min

Advanced

Spectrum

Store: None

Range from: 190.0 to 400.0 nm

Step: 2.0 nm

Threshold: 10.0 mAU

Analog Output

Output 1:

Zero Offset: 5 %

Attenuation: 1000 mAU

Margin for negative Absorbance

100 mAU

Slit

4 nm

Autobalance

☒ Prerun

☐ Postrun

Lamps on required for acquisition

☒ UV Lamp

Timetable (empty)

☐ Show timetable graph

OK Apply Cancel Help

Figure 12 Method parameter settings

NOTE

For additional help and support: Highlight the desired cell and press **F1**. A help screen will open with additional information and documentation about the topic.

3 Using the Module

Method Parameter Settings

Table 4 Method Parameter Settings

Signals						
	Use Signal	Wave length	Band width	Reference Wavelength	Reference Bandwidth	
Signal A	☒	254.0	4.0	☒	360.0	100.0 nm
Signal B	☒	210.0	4.0	☒	360.0	100.0 nm
Signal C	☒	214.0	4.0	☒	360.0	100.0 nm
Signal D	☒	230.0	4.0	☒	360.0	100.0 nm
Signal E	☒	260.0	4.0	☒	360.0	100.0 nm
Signal F	☐	273.0	4.0	☒	360.0	100.0 nm
Signal G	☐	280.0	4.0	☒	360.0	100.0 nm
Signal H	☐	250.0	4.0	☒	360.0	100.0 nm

Peakwidth

> 0.1 min (2 s response time) (2.5 Hz)

Stoptime

As P

min

< 0.0008 min (0.008 s response time) (240 Hz)
 > 0.0008 min (0.016 s response time) (240 Hz)
 > 0.0016 min (0.031 s response time) (160 Hz)
 > 0.0031 min (0.063 s response time) (80 Hz)
 > 0.0063 min (0.13 s response time) (40 Hz)
 > 0.013 min (0.25 s response time) (20 Hz)
 > 0.025 min (0.5 s response time) (10 Hz)
 > 0.05 min (1 s response time) (5 Hz)
 > 0.1 min (2 s response time) (2.5 Hz)
 > 0.2 min (4 s response time) (1.25 Hz)
 > 0.4 min (8 s response time) (0.62 Hz)
 > 0.85 min (16 s response time) (0.31 Hz)

Figure 13 G7117B Peakwidth Settings

Peakwidth

> 0.1 min (2 s response time) (2.5 Hz)

Stoptime

As P

min

< 0.0016 min (0.016 s response time) (120 Hz)
 > 0.0016 min (0.031 s response time) (120 Hz)
 > 0.0031 min (0.063 s response time) (80 Hz)
 > 0.0063 min (0.13 s response time) (40 Hz)
 > 0.013 min (0.25 s response time) (20 Hz)
 > 0.025 min (0.5 s response time) (10 Hz)
 > 0.05 min (1 s response time) (5 Hz)
 > 0.1 min (2 s response time) (2.5 Hz)
 > 0.2 min (4 s response time) (1.25 Hz)
 > 0.4 min (8 s response time) (0.62 Hz)
 > 0.85 min (16 s response time) (0.31 Hz)

Figure 14 G7117A Peakwidth Settings

Signals

Up to 8 individual signals can be set. For each of the signals, the wavelength and bandwidth can be set for sample and reference.

Limits:

- Wavelength: 190.0 to 640.0 nm in steps of 0.1 nm
- Bandwidth: 1.0 to 400.0 nm in steps of 0.1 nm

Setting an appropriate reference wavelength could improve the baseline behavior.

Peakwidth (Responsetime, Data Rate)

Peakwidth enables you to select the peak width (response time) for your analysis. The peak width is defined as the width of a peak, in minutes, at half the peak height. Set the peak width to the narrowest expected peak in your chromatogram. The peak width sets the optimum response time for your detector. The peak detector ignores any peaks that are considerably narrower, or wider, than the peak width setting. The response time is the time between 10 % and 90 % of the output signal in response to an input step function. When the All spectrum storage option is selected, then spectra are acquired continuously depending on the setting of the peak width. The time specified by the peak width is used as a factor in the acquisition of spectra. The acquisition time for one spectrum is slightly less than the peak width divided by 8, which is the acquisition time.

Limits: When you set the peak width (in minutes), the corresponding response time is set automatically and the appropriate data rate for signal and spectra acquisition is selected.

Do not use peak width shorter than necessary.

G7117A: Do not use 0.025 s response time (no filtering/high noise and no need (actually ultra-fast LC doesn't deliver peaks <0.0025 min/<0.15 s).

NOTE

The 1290 Infinity II DAD FS (G7117A) has a data rate of up to 120 Hz. The 1290 Infinity II DAD (G7117B) has a data rate of up to 240 Hz.

For details see "Peak width (response time)" on page 73.

Table 4 Method Parameter Settings

Peakwidth (time programmed)

These selections can be made during time programmed operation. When used in a timetable, Peakwidth changes the filters used for peak-controlled spectra acquisition, but not the data rate of a chromatographic signal.

NOTE

This setting makes sense only with peak-controlled spectra; it allows you to change the peakwidth setting to account for broadening peaks at the end of the run.

Stoptime/Posttime

The stoptime is the time where either the complete system stops (**As Pump/Injector**) or the module (if different from system stop time). The data collection is stopped at this time. A posttime period can be used to allow module's items to equilibrate (e.g. after gradient change or temperature change).

Spectrum Settings

Analog Output

The range can be set to either 100 mV or 1 V full scale, see "[Detector Control Settings](#)" on page 42.

- **Zero Offset:** 1 – 99 % in steps of 1 % (5 % equal to 50 mV).
- **Attenuation:** 0.98 – 2000 mAU at discrete values for either 100 mV or 1 V full scale.

3 Using the Module
Method Parameter Settings

Table 4 Method Parameter Settings

Margin for negative Absorbance	Slit
100 mAU	4 nm

Margin for negative Absorbance

Use this field to modify the detector's signal handling to increase the margin for negative absorbance. Use this option if, for example, your solvent gradient produces a decreasing baseline absorbance, and for GPC analyses. Limits: 100 to 4000 mAU.

The higher the value the greater the baseline noise. Set this value only if you expect negative absorbance greater than -100 mAU.

Slitwidth (G7117B): You can select the optical bandwidth (1, 2, 4 or 8 nm) of the detector; the narrower the slit, the smaller the optical bandwidth of the instrument, but the lower its sensitivity. The smaller the optical bandwidth the higher the spectral resolution.

Slitwidth (G7117A): The 1290 Infinity DAD FS (G7117A) has a fixed slit width of 4 nm.

Autobalance	Lamps on required for acquisition
☒ Prerun ☐ Postrun	☒ UV Lamp

Autobalance

Defines, whether a balance is performed prior to a run and/or after a run has finished.

Lamp on required for acquisition:

If unchecked, the lamp will be turned off after the analysis has finished. Note that the lamp on requires at least one hour warm-up time, see ["Warm up of the Detector"](#) on page 94.

▼ Timetable (1/100 events)		
Time	Function	Parameter
0	Change Peakdetector P...	Peakthreshold: same as Peakwidth
Balance Change Signal Change Threshold Change Peakdetector Peakwidth Change Spectra Acquisition Mode		
Add Remove Clear all Cut Copy Paste		



Timetable

You may set up time events to change functions with their parameters over the run time. Add lines as required.

Time Limits: 0.00 to 99999.00 min in steps of 0.01 min.

Via the buttons in the bottom area, time table lines can be added, removed, cut copied, pasted or completely cleared. Based on the chosen function, a certain parameter can be selected.

Table 4 Method Parameter Settings



Instrument Curves


The detector has several signals (internal temperatures, voltages of lamps) that can be used for diagnosing problems. These can be baseline problems deriving from deuterium lamps wander / drift problems due to temperature changes.

These signals can be used in addition to the normal baseline signal to determine whether correlation to temperature or voltage/current of the lamp.

These signals are available via the Agilent ChemStation Online Plot/Data Signal and/or Agilent Lab Advisor Software.

3 Using the Module

Method Parameter Settings



4 Preparing the Module

Leak and Waste Handling	50
Waste Concept	51
Setting up an Analysis	52
Before Using the System	52
Requirements and Conditions	54
Preparing the Detector	56
Preparing the HPLC System	56
Running the Sample and Verifying the Results	58
Solvent Information	59

This chapter provides information on how to set up the module for an analysis and explains the basic settings.



Leak and Waste Handling

WARNING

Toxic, flammable and hazardous solvents, samples and reagents

The handling of solvents, samples and reagents can hold health and safety risks.

- When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.
- The volume of substances should be reduced to the minimum required for the analysis.
- Do not operate the instrument in an explosive atmosphere.
- Never exceed the maximal permissible volume of solvents (6 L) in the solvent cabinet.
- Do not use bottles that exceed the maximum permissible volume as specified in the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.
- Arrange the bottles as specified in the usage guideline for the solvent cabinet.
- A printed copy of the guideline has been shipped with the solvent cabinet, electronic copies are available on the Internet.
- Ground the waste container.
- The residual free volume in the appropriate waste container must be large enough to collect the waste liquid.
- Check the filling level of the waste container regularly.
- To achieve maximal safety, check the correct installation regularly.
- Do not use solvents with an auto-ignition temperature below 200 °C (392 °F).

NOTE

Recommendations for Solvent Cabinet

For details, see the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.

For correct installation of your system contact your Agilent service representative.

Waste Concept

- 1 Agilent recommends using the 6 L waste can with 1 Stay Safe cap GL45 with 4 ports (5043-1221) for optimal and safe waste disposal. If you decide to use your own waste solution, make sure that the tubes don't immerse in the liquid.



Setting up an Analysis

This chapter can be used for

- preparing the system,
- to learn the set up of an HPLC analysis and
- to use it as an instrument check to demonstrate that all modules of the system are correctly installed and connected. It is not a test of the instrument performance.
- Learn about special settings

Before Using the System

Solvent Information

Observe recommendations on the use of solvents in chapter “Solvents” in the pump’s reference manual.

Priming and Purging the System

When the solvents have been exchanged or the pumping system has been turned off for a certain time (for example, overnight) oxygen will re-diffuse into the solvent channel between the solvent reservoir, vacuum degasser (when available in the system) and the pump. Solvents containing volatile ingredients will slightly lose these. Therefore priming of the pumping system is required before starting an application.

Table 5 Choice of Priming Solvents for Different Purposes

Activity	Solvent	Comments
After an installation	Isopropanol	Best solvent to flush air out of the system
When switching between reverse phase and normal phase (both times)	Isopropanol	Best solvent to flush air out of the system
After an installation	Ethanol or Methanol	Alternative to Isopropanol (second choice) if no Isopropanol is available
To clean the system when using buffers	Bidistilled water	Best solvent to re-dissolve buffer crystals
After a solvent change	Bidistilled water	Best solvent to re-dissolve buffer crystals
After the installation of normal phase seals (P/N 0905-1420)	Hexane + 5% Isopropanol	Good wetting properties

NOTE

The pump should never be used for priming empty tubings (never let the pump run dry). Use a syringe to draw enough solvent for completely filling the tubings to the pump inlet before continuing to prime with the pump.

- 1 Open the purge valve of your pump (by turning it counterclockwise) and set flow rate to 3 – 5 mL/min.
- 2 Flush all tubes with at least 30 mL of solvent.
- 3 Set flow to required value of your application and close the purge valve.

NOTE

Pump for approximately 10minutes before starting your application.

Requirements and Conditions

What You Will Need

The table below lists the items you need to have for the set up of the analysis. Some of these are optional (not required for the basic system).

Table 6 What you will need

Agilent 1200 Infinity Series system	Pump (plus degassing)
	Autosampler
	Detector, standard flow cell installed
	Degasser (optional)
	Column Compartment (optional)
	Agilent CDS
	System should be correctly set up for LAN communication with the Agilent ChemStation
Column:	Zorbax Eclipse XDB-C18, 4.6 x 150 mm, 5 µm (993967-902) or an equivalent column
Standard:	Agilent isocratic checkout sample (01080-68704)

Conditions

A single injection of the isocratic test standard is made under the conditions given in [Table 7](#) on page 54:

Table 7 Conditions

Flow	1.5 mL/min
Stoptime	8 min
Solvent	100% (30% water/70% Acetonitrile)
Temperature	Ambient
Wavelength	sample 254 nm
Injection Volume	1 µL
Column Temperature (optional):	25 °C or ambient

Typical Chromatogram

A typical chromatogram for this analysis is shown in [Figure 15](#) on page 55. The exact profile of the chromatogram will depend on the chromatographic conditions. Variations in solvent quality, column packing, standard concentration and column temperature will all have a potential effect on peak retention and response.




Figure 15 Typical Chromatogram with UV-detector

Preparing the Detector

For best performance of the detector

- Let the lamp warm-up and stabilize for at least one hour (initial turn on of the module requires a longer time depending on the environment and the application needs); refer to “[Specification Conditions](#)” on page 32.
- For high sensitivity measurements, a stable environment is required; refer to “[Environment](#)” on page 26. Prevent drafts from air condition systems.
- Setting an appropriate reference wavelength could improve the baseline behavior.
- Do not work with removed/open front panels/doors. When the system includes a G1316 TCC (typically located below the detector) and its front panel is removed while the TCC is set to high temperatures, the up-streaming air could influence the stability of the detector baseline.

Preparing the HPLC System

- 1 Turn on the control software and the monitor.
- 2 Turn on the modules.
- 3 Start the control software. The screen should show all modules and the system status is Not Ready.
- 4 Turn on the modules that require conditioning:
 - a Detector lamp (warm-up for at least 60 min to get a stable baseline).
 - b Column compartment (set temperature as required).
 - c Pump (purge).
 - d Sampler (prepare the standard isocratic sample into a vial).
 - e Solvents (fill water and Acetonitrile into the solvent bottles).
- 5 Load the default method.
- 6 Pump the water/acetonitrile (30/70 %) mobile phase through the column for 10 min for equilibration.
- 7 Select the menu item **Run Control > Sample Info** and enter information about this application. Click **OK** to leave this screen.

Sample Info: WADI1171

Operator name: SYSTEM

Data file

Path: D:\Agilent\ChemStation\DATA\ Subdirectory: G7117

Name Pattern

Signal 1: <SampleName> <Date> <Time> x Isocratic Sample 2014-08-27 10-42-06.D

Sample parameters

Vial/Location: 1 (blank run if no entry)

Sample name: Isocratic Sample Sample amount: 0

Multiplier: 1 Dilution: 1 ISTD amount: 0

Comment: Isocratic Sample

Custom Fields ... Run Method OK Cancel Help

Figure 16 Sample Info

Running the Sample and Verifying the Results

- 1 To start a run select the menu item **RunControl > Run Method**.
- 2 This will start the modules and the online plot on the Agilent ChemStation will show the resulting chromatogram.




Figure 17 Chromatogram with Isocratic Test Sample

NOTE

Information about using the Data Analysis functions can be obtained from the Using your ChemStation manual supplied with your system.

Solvent Information

Observe the following recommendations on the use of solvents.

- Follow recommendations for avoiding the growth of algae, see pump manuals.
- Small particles can permanently block capillaries and valves. Therefore, always filter solvents through 0.4 µm filters.
- Avoid or minimize the use of solvents that may corrode parts in the flow path. Consider specifications for the pH range given for different materials like flow cells, valve materials etc. and recommendations in subsequent sections.

Material Information

Materials in the flow path are carefully selected based on Agilent's experiences in developing highest quality instruments for HPLC analysis over several decades. These materials exhibit excellent robustness under typical HPLC conditions. For any special conditions, please consult the material information section or contact Agilent.

Disclaimer

Subsequent data were collected from external resources and are meant as a reference. Agilent cannot guarantee the correctness and completeness of such information. Data is based on compatibility libraries, which are not specific for estimating the long-term life time under specific but highly variable conditions of UHPLC systems, solvents, solvent mixtures and samples. Information can also not be generalized due to catalytic effects of impurities like metal ions, complexing agents, oxygen etc. Apart from pure chemical corrosion, other effects like electro corrosion, electrostatic charging (especially for non-conductive organic solvents), swelling of polymer parts etc. need to be considered. Most data available refers to room temperature (typically 20 – 25 °C, 68 – 77 °F). If corrosion is possible, it usually accelerates at higher temperatures. If in doubt, please consult technical literature on chemical compatibility of materials.

PEEK

PEEK (Polyether-Ether Ketones) combines excellent properties regarding biocompatibility, chemical resistance, mechanical and thermal stability. PEEK is therefore the material of choice for UHPLC and biochemical instrumentation.

There is still a number of known incompatibilities with chemicals such as chloroform, methylene chloride, THF, DMSO, strong acids (nitric acid > 10 %, sulphuric acid > 10 %, sulfonic acids, trichloroacetic acid), halogenes or aqueous halogene solutions, phenol and derivatives (cresols, salicylic acid etc.).

Polyimide

Agilent uses semi-crystalline polyimide for rotor seals in valves and needle seats in autosamplers. One supplier of polyimide is DuPont, which brands polyimide as Vespel, which is also used by Agilent.

Polyimide is stable in a pH range between 1 and 10 and in most organic solvents. It is incompatible with concentrated mineral acids (e.g. sulphuric acid), glacial acetic acid, DMSO and THF. It is also degraded by nucleophilic substances like ammonia (e.g. ammonium salts in basic conditions) or acetates.

Polyethylene (PE)

Agilent uses UHMW (ultra-high molecular weight)-PE/PTFE blends for yellow piston and wash seals, which are used in 1290 Infinity pumps and for normal phase applications in 1260 Infinity pumps.

Polyethylene has a good stability for most common inorganic solvents including acids and bases in a pH range of 1 to 12.5. It is compatible to many organic solvents used in chromatographic systems like methanol, acetonitrile and isopropanol. It has limited stability with aliphatic, aromatic and halogenated hydrocarbons, THF, phenol and derivatives, concentrated acids and bases. For normal phase applications, the maximum pressure should be limited to 200 bar.


Tantalum (Ta)

Tantalum is inert to most common HPLC solvents and almost all acids except fluoric acid and acids with free sulfur trioxide. It can be corroded by strong bases (e.g. hydroxide solutions > 10 %, diethylamine). It is not recommended for the use with fluoric acid and fluorides.

Stainless Steel (ST)

Stainless steel is inert against many common solvents. It is stable in the presence of acids and bases in a pH range of 1 to 12.5. It can be corroded by acids below pH 2.3. It can also corrode in following solvents:

- Solutions of alkali halides, their respective acids (for example, lithium iodide, potassium chloride, and so on) and aqueous solutions of halogens.
- High concentrations of inorganic acids like nitric acid, sulfuric acid and organic solvents especially at higher temperatures (replace, if your chromatography method allows, by phosphoric acid or phosphate buffer which are less corrosive against stainless steel).
- Halogenated solvents or mixtures which form radicals and/or acids, for example:



This reaction, in which stainless steel probably acts as a catalyst, occurs quickly with dried chloroform if the drying process removes the stabilizing alcohol.

- Chromatographic grade ethers, which can contain peroxides (for example, THF, dioxane, di-isopropylether). Such ethers should be filtered through dry aluminium oxide which adsorbs the peroxides.
- Solutions of organic acids (acetic acid, formic acid, and so on) in organic solvents. For example, a 1 % solution of acetic acid in methanol will attack steel.
- Solutions containing strong complexing agents (for example, EDTA, ethylene diamine tetra-acetic acid).
- Mixtures of carbon tetrachloride with 2-propanol or THF.

Diamond-Like Carbon (DLC)

Diamond-Like Carbon is inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

Fused silica and Quartz (SiO₂)

Fused silica is used in 1290 Infinity Flow Cells and capillaries. Quartz is used for classical flow cell windows. It is inert against all common solvents and acids except hydrofluoric acid and acidic solvents containing fluorides. It is corroded by strong bases and should not be used above pH 12 at room temperature. The corrosion of flow cell windows can negatively affect measurement results. For a pH greater than 12, the use of flow cells with sapphire windows is recommended.

Gold

Gold is inert to all common HPLC solvents, acids and bases within the specified pH range. It can be corroded by complexing cyanides and concentrated acids like aqua regia.

Zirconium Oxide (ZrO₂)

Zirconium Oxide is inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

Platinum/Iridium

Platinum/Iridium is inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

Fluorinated polymers (PTFE, PFA, FEP, FFKM)

Fluorinated polymers like PTFE (polytetrafluorethylene), PFA (perfluoroalkoxy) and FEP (fluorinated ethylene propylene) are inert to almost all common acids, bases, and solvents. FFKM is perfluorinated rubber, which is also resistant to most chemicals. As an elastomer, it may swell in some organic solvents like halogenated hydrocarbons.

TFE/PDD copolymer tubings, which are used in all Agilent degassers except G1322A, are not compatible with fluorinated solvents like Freon, Fluorinert, or Vertrel. They have limited life time in the presence of


Hexafluoroisopropanol (HFIP). To ensure the longest possible life with HFIP, it is best to dedicate a particular chamber to this solvent, not to switch solvents, and not to let dry out the chamber. For optimizing the life of the pressure sensor, do not leave HFIP in the chamber when the unit is off.

Sapphire, Ruby and Al₂O₃-based ceramics

Sapphire, ruby and ceramics based on aluminum oxide Al₂O₃ are inert to almost all common acids, bases and solvents. There are no documented incompatibilities for HPLC applications.

4 **Preparing the Module**

Solvent Information



5 Optimizing the Detector

Introduction	66
Optimization Overview	67
Choosing a Flow Cell	68
Inline Pressure Relief Valve Kit (G4212-68001)	70
Application Information	70
Special Information of 60 mm Cartridge Flow Cell	71
Optimizing for Sensitivity, Selectivity, Linearity and Dispersion	72
Flow Cell Path Length	72
Peak width (response time)	73
Sample and Reference Wavelength and Bandwidth	77
Slit Width (G7117B)	79
Optimizing Spectral Acquisition	81
Margin for Negative Absorbance	82
Optimizing Selectivity	83
Quantifying Coeluting Peaks by Peak Suppression	83
Ratio Qualifiers for Selective Detection of Compound Classes	85
Optimizing the Detector Regarding to the System	87
Delay Volume and Extra-Column Volume	87
How to Configure the Optimum Delay Volume	87
How to Achieve Higher Sensitivity	88
Warm up of the Detector	94

This chapter provides information on how to optimize the detector.



Introduction

The detector has a variety of parameters that can be used to optimize performance. Depending on whether signal or spectral data need to be optimized, different settings are recommended. The following sections describe optimization for:

- signal sensitivity, selectivity and linearity,
- spectral sensitivity and resolution (DAD only), and
- disk space required for storing data.

NOTE

The information in this chapter should be seen as a basic introduction to diode array detector techniques. Some of these techniques may not be available in the instrument software controlling the detector.

How to Get the Best Detector Performance

The information below will guide you on how to get the best detector performance. Follow these rules as a start for new applications. It gives rules-of-thumb for optimizing detector parameters.

Optimization Overview

Table 8 Optimization Overview

Parameter	Impact
1 Selection of flow cell Choose flow cell according to used column ("Choosing a Flow Cell" on page 68).	peak resolution versus sensitivity
2 Connection of flow cell	chromatographic resolution
3 Setting the peak width (response time) - Use peak width according "Choosing a Flow Cell" on page 68 as starting point. - Set the peak-width close to the width of a narrow peak of interest in your chromatogram.	peak resolution versus sensitivity versus disk space
4 Setting wavelength and bandwidth - Sample wavelength: - Never miss a peak by the use of a browser wavelength like 250 nm with 100 nm bandwidth. - Select specific wavelength with reduced optical bandwidth if you need selectivity, e.g. 254.0 nm / 4 nm and 360.0 nm / 100 nm as reference wavelength. - Set the sample wavelength to a peak or valley to get best linearity in general; select a valley to get best linearity for high concentrations. - Reference wavelength: - Select the reference wavelength with broad bandwidth (30...100 nm) wavelength range where your analytes have little or no absorbance (e.g. sample at 254 nm, reference at 320 nm). - Select the reference wavelength as near as possible to the UV range.	- sensitivity versus selectivity - sensitivity versus linearity - baseline drift due to RI effects.

Table 8 Optimization Overview

Parameter	Impact
5 Setting the slit width (G7117B only)	
• Use 4 nm slit for normal applications. • Use narrow slit (e.g 1 nm) if your analytes have narrow absorbance bands and for high concentrations. • Use a wide slit (e.g. 8 nm) to detect very low concentrations. • Optimizing spectral acquisition (DAD only) • Set the spectral wavelength range (for colorless samples 190...400 nm is sufficient). • Set step to 4 nm for normal use; set small step (and slit width) if high resolution of spectra with fine structure is wanted.	• spectral resolution, sensitivity and linearity.

Choosing a Flow Cell

Several flavors of the Max-Light Cartridge Flow Cell are available, see [Table 9](#) on page 68.

Table 9 Specifications for Max-Light Cartridge Flow Cells

Cartridge Cells	• Max-Light Cartridge Cell (10 mm, V(s) 1.0 µL) (G4212-60008) • Max-Light Cartridge Cell (60 mm, V(s) 4.0 µL) (G4212-60007) • HDR Max-Light Cartridge Cell (3.7 mm, V(s) 0.4 µL) (G4212-60032) • ULD Max-Light Cartridge Cell (10 mm, V(s) 0.6 µL) (G4212-60038) • Max-Light Cartridge Test Cell (G4212-60011)
Maximum pressure	70 bar (1015 psi) Maximum Operating Pressure (MOP) ¹ 150 bar (2175 psi) Maximum Incidental Pressure (MIP) ²
pH range	1.0-12.5 (solvent dependent)

¹ Maximum Operating Pressure (MOP): The maximum pressure at which the system can operate continuously under normal conditions.

² Maximum Incidental Pressure (MIP): The maximum pressure which the system can experience during a short time.

High Sensitivity

If higher sensitivity is necessary, the Max-Light Cartridge Cell (60 mm, V(s) 4.0 μ L) (G4212-60007) can be used. This cell enhances the detector by lowering the limit of detection (LOD) by a factor of about 3 (depending on the application).

Normal Applications

The Max-Light Cartridge Cell (10 mm, V(s) 1.0 μ L) (G4212-60008) covers a wide range of applications:

- all column diameter down to at least 2.1 mm ID or even less
- applications with peak dispersion (Peakwidth x flow) down to $\sim 2 \mu$ L
[example: pw = 0.04 min at flow = 0.1 mL/min gives peak dispersion of
 $0.04 \text{ min} \times 0.1 \text{ mL/min} = 0.004 \text{ mL} = 4 \mu\text{L}$]

Ultra-Low Dispersion

The Max-Light Cartridge ULD cell can be used with the G7117A DAD FS and G7117B DAD. The cell is a requirement for the Ultra-Low Dispersion Kit solution which currently exists as 1290 Infinity Ultra-Low Dispersion Kit (5067-5189). The cell should be part of the ultra-low dispersion solution.

High Dynamic Range

The Max-Light Cartridge HDR cell can be used with the G7117A DAD FS and G7117B DAD. The cell is required as a part of the High Dynamic Range (HDR) solution.

NOTE

To protect the flow cell against overpressure (e.g. in systems with LC/MS) install Inline Pressure Relief Valve Kit (G4212-68001), see ["Inline Pressure Relief Valve Kit \(G4212-68001\)"](#) on page 70.

Recommendations

For G4212-60007 and G4212-60008

The use of Peek-FS capillaries is not recommended. In combination with the SST zero dead volume fitting (e.g. at the inlet) the capillary could break and the glass particles could block/damage the flow cell.

Inline Pressure Relief Valve Kit (G4212-68001)

When several detectors are installed in a system the connecting capillary and fittings between the detectors must be carefully chosen to keep chromatographic influence on peak shape small. On the other hand narrow bore connection capillaries generate a significant pressure drop dependent on flow rate and solvent properties.

The pressure relief valve is designed to protect the flow cell of a Agilent 1200 Series Infinity Diode Array Detector (G7117A DAD FS and G7117B DAD). Agilent strongly recommends installing the pressure relief valve at the outlet of the detector as soon as a second detector is installed like in LC/MS applications.

The pressure relief valve with a low internal volume check valve. The dead volume is smaller than 100 nL delay volume (inlet to outlet). The ball of the check valve is spring loaded and adjusted to open at typically 100 bar. On overpressure (typically around 100 bar) it releases the pressure to waste.

Application Information

For the analysis and characterization of proteins and large biomolecules for SEC, AEX and RP applications add 100 mM salt into mobile phase or 10 % organic to prevent secondary interaction.

For cation exchange chromatography the usage of an Agilent Diode Array Detector G1315C/D with the respective bio-inert flow cell is highly recommended to avoid unspecific interaction of the protein with the flow cell.

For applications with mobile phases of a pH above 12.5 use an Agilent Diode Array Detector G1315C/D and the respective bio-inert flow cell.

Special Information of 60 mm Cartridge Flow Cell

Application Information

The geometrical volume of the 60 mm cell is 6 times larger than the 10 mm cell. However, the chromatographic relevant dispersion volume, the square roots of variances, accounting for cell specific geometrical volume shape and fluidic flow pattern, have been determined as $\sigma V = 4 \mu\text{L}$ and $\sigma V = 1 \mu\text{L}$ in for the 10 mm cell.

Due to the larger dispersion volume, the 60 mm cell is primarily designed for 4.6 mm column applications to achieve highest sensitivity with no additional peak broadening. However, if sensitivity is important the 60 mm cell will also be advantageous in case of smaller columns (3 mm, 2.1 mm) but depending on the chromatographic system and method additional peak broadening might occur.

The upper limit of concentration

Care should be taken in methods where high background absorption of solvents or modifiers are present. When using the 60 mm cell the detector will measure 6 times the background absorption as in case of the 10 mm cell, which will reduce the remaining dynamic absorbance range for sample peaks. Furthermore those UV absorbing modifiers could compromise the sensitivity gain (signal/noise) of 60 mm cell.

The linearity limit of the detector is seen at about 2 AU for both, the 10 mm and the 60 mm Max-Light Cartridge Flow Cell. Using firmware revision B.06.25 and below, the 60 mm Max-Light Cartridge Cell linearity limit would be 333 mAU/cm.

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion

Flow Cell Path Length

Lambert-Beer's law shows a linear relationship between the flow cell path length and absorbance.

$$\text{Absorbance} = -\log T = \log \frac{I_0}{I} = \epsilon \times C \times d$$

where

T is the transmission, defined as the quotient of the intensity of the transmitted light I divided by the intensity of the incident light, I_0 ,

ϵ is the extinction coefficient, which is a characteristic of a given substance under a precisely-defined set of conditions of wavelength, solvent, temperature and other parameters,

C [mol/L] is the concentration of the absorbing species, and

d [m] is the path length of the cell used for the measurement.

The detector can now output the signal in two forms:

- 1 In Absorbance divide by the path length AU/cm, that is then similar to [$\epsilon \times C$]. Advantage: samples with same concentration have same peak height also at cells with different path lengths.

The upper limit of concentration: the linearity limit of the detector is then seen at about 2 AU/path length, so for the 6 cm Max-Light Cartridge Cell the linearity limit is 333 mAU/cm].

- 2 In AU that is equal to $\epsilon \times C \times d$ like normal done in the past: now for recalculation to your concentration C the path length must be considered.

Therefore, flow cells with longer path lengths yield higher signals. Although noise usually increases little with increasing path length, there is a gain in signal-to-noise ratio.

When increasing the path length, the cell volume could increase. Depending on the peak volume, this could cause more peak dispersion.

As a rule-of-thumb the flow cell volume should be about 1/3 of the peak volume at half height. To determine the volume of your peaks, take the peak width as reported in the integration results multiply it by the flow rate and divide it by 3).

NOTE

This may result in problems when the used peak width is set to large and all peaks are filtered accordingly.

Traditionally LC analysis with UV detectors is based on comparing measurements with internal or external standards. To check photometric accuracy of the Agilent detector it is necessary to have more precise information on path lengths of the detector flow cells.

Part Number	Path Length	Cell Volume (σ)
G4212-60008/G5615-60018	1.0 cm	1.0 μ L
G4212-60007/G5615-60017	6.0 cm	4.0 μ L

Peak width (response time)

Response time describes how fast the detector signal follows a sudden change of absorbance in the flow cell. The detector uses digital filters to adapt response time to the width of the peaks in your chromatogram. These filters do not affect peak area nor peak symmetry. When set correctly, such filters reduce baseline noise significantly ([Figure 18](#) on page 74), but reduce peak height only slightly. In addition, these filters reduce the data rate to allow optimum integration and display of your peaks and to minimize disk space required to store chromatograms and spectra.

5 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion




Figure 18 Influence of Response Time on Signal and Noise

Table 10 on page 75 and Table 11 on page 76 list the filter choices of the detector. To get optimum results, set peak width as close as possible to a narrow peak of interest in your chromatogram. Response time will then be approximately 1/3 of the peak width, resulting in less than 5 % peak-height reduction and less than 5 % additional peak dispersion. Decreasing the peak width setting in the detector will result in less than 5 % gain in peak height but baseline noise will increase by a factor of 1.4 for a factor of 2 response-time reduction. Increasing the peak width (response time) by a factor of two from the recommended setting (over-filtering) will reduce peak height by about 20 % and reduce baseline noise by a factor of 1.4. This gives you the best possible signal-to-noise ratio, but may affect peak resolution.

Table 10 Peak Width — Response Time — Data Rate (G7117B)

Peak width at half height [min] ¹	Response [s]	Signal data rate [Hz]	Scan data rate [HZ] ≤126 pts/scan	Scan data rate [HZ] ≤251 pts/scan	Scan data rate [HZ] ≤501 pts/scan	Scan data rate [HZ] >501 pts/scan
< 0.00078125	0.0078125	240	240	80	40	20
> 0.00078125	0.015625	240	240	80	40	20
> 0.0015625	0.03125	160	160	80	40	20
> 0.003125	0.0625	80	80	80	40	20
> 0.00625	0.125	40	40	40	40	20
> 0.0125	0.25	20	20	20	20	20
> 0.025	0.5	10	10	10	10	10
> 0.05	1	5	5	5	5	5
> 0.1	2	2.5	2.5	2.5	2.5	2.5
> 0.2	4	1.25	1.25	1.25	1.25	1.25
> 0.4	8	0.625	0.625	0.625	0.625	0.625
> 0.85	16	0.3125	0.3125	0.3125	0.3125	0.3125

¹ Values in the User Interface may be rounded.

5 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion

Table 11 Peak Width — Response Time — Data Rate (G7117A)

	Peak width at half height [min] ¹	Response [s]	Scan data rate[Hz] ≤251 pts/scan	Scan data rate[Hz] ≤501 pts/scan	Scan data rate[Hz] > 501 pts/scan
< 0.0015625	0.015625	120	120	40	20
> 0.0015625	0.03125	120	120	40	20
> 0.003125	0.0625	80	80	40	20
> 0.00625	0.125	40	40	40	20
> 0.0125	0.25	20	20	20	20
> 0.025	0.5	10	10	10	10
> 0.05	1	5	5	5	5
> 0.1	2	2.5	2.5	2.5	2.5
> 0.2	4	1.25	1.25	1.25	1.25
> 0.4	8	0.625	0.625	0.625	0.625
> 0.85	16	0.3125	0.3125	0.3125	0.3125

¹ Values in the User Interface may be rounded.

NOTE

The maximum spectra scan rate depends on the data points per scan, see [Table 10](#) on page 75 and [Table 11](#) on page 76. Running at 160 Hz, the spectra scan data rate is reduced automatically if the spectra scan data rate is more than 251 points/scan.

Sample and Reference Wavelength and Bandwidth

The detector measures absorbance simultaneously at wavelengths from 190 to 640 nm. A UV-lamp provides good sensitivity over the whole wavelength range.

If you know little about the analytes in your sample, store all spectra over the full wavelength range. This provides full information but fills up your disk space rather quickly. Spectra can be used to check a peak's purity and identity. Spectral information is also useful to optimize wavelength settings for your chromatographic signal.

The detector can compute and store at run time up to 8 signals with these properties:

- sample wavelength, the center of a wavelength band with the width of sample bandwidth (BW), and optionally
- reference wavelength, the center of a wavelength band with the width of reference bandwidth.

The signals comprises a series of data points over time, with the average absorbance in the sample wavelength band minus the average absorbance of the reference wavelength band.

Signal A in the detector default method is set to sample 254.0/4, reference 360.0/100, that is, the average absorbance from 252 – 256 nm minus the average absorbance from 310 – 410 nm. As all analytes show higher absorbance at 252 – 256 nm than at 310 – 410 nm, this signal will show you virtually every compound which can be detected by UV absorbance.

Many compounds show absorbance bands in the spectrum. [Figure 19](#) on page 78 shows the spectrum of anisic acid as an example. To optimize for lowest possible detectable concentrations of anisic acid, set the sample wavelength to the peak of the absorbance band (that is, 252 nm) and the sample bandwidth to the width of the absorbance band (that is, 30 nm). A reference of 360,100 is adequate. Anisic acid does not absorb in this range.

If you work with high concentrations, you may get better linearity above 1.5 AU by setting the sample wavelength to a valley in the spectrum, like 225 nm for anisic acid.

5 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion




Figure 19 Optimization of Wavelength Setting

A wide bandwidth has the advantage of reducing noise by averaging over a wavelength range — compared to a 4 nm bandwidth, the baseline noise is reduced by a factor of approximately 2.5, whereas the signal is about 75 % of a 4 nm wide band. The signal-to-noise ratio for a 30 nm bandwidth is twice that for a 4 nm bandwidth in our example.




Figure 20 Influence of Bandwidth on Signal and Noise

Because the detector averages absorbance values that are calculated for each wavelength, using a wide bandwidth does not negatively impact linearity.

The use of a reference wavelength is highly recommended to further reduce baseline drift and wander induced by room temperature fluctuations or refractive index changes during a gradient.

An example of the reduction of baseline drifts is shown in [Figure 21](#) on page 79 for PTH-amino acids. Without a reference wavelength, the chromatogram drifts downwards due to refractive index changes induced by the gradient. This is almost completely eliminated by using a reference wavelength. With this technique, PTH-amino acids can be quantified in the low picomole range even in a gradient analysis.




Figure 21 Gradient Analysis of PTH-Amino Acids (1 pmol each), with and without Reference

Slit Width (G7117B)

The 1290 Infinity DAD (G7117B) has a variable slit at the entrance of the spectrograph. This is an effective tool to adapt the detector to changing demand of different analytical problems.

5 Optimizing the Detector

Optimizing for Sensitivity, Selectivity, Linearity and Dispersion

A narrow slit provides spectral resolution for analytes with very fine structures in the absorbance spectrum. An example of such a spectrum is benzene. The five main absorbance bands (fingers) are only 2.5 nm wide and just 6 nm apart from each other.






Figure 22 Benzene at 1 and 4 nm slit width (principle)

A wide slit uses more of the light shining through the flow cell. This gives lower baseline noise as shown in [Figure 23](#) on page 80.

Slit width = 1 nm



Slit width = 8 nm



Data Rate = 160 Hz

Figure 23 Influence of the Slit Width on Baseline Noise

However, with a wider slit, the spectrograph's optical resolution (its ability to distinguish between different wavelengths) diminishes. Any photodiode receives light within a range of wavelength determined by the slit width. This explains why the fine spectral structure of benzene disappears when using a 8 nm wide slit.

Furthermore, the absorbance is no longer strictly linear with concentration for wavelengths at a steep slope of a compound's spectrum.

Substances with fine structures and steep slopes like benzene are very rare.

In most cases the width of absorbance bands in the spectrum is more like 30 nm as with anisic acid ([Figure 19](#) on page 78).

In most situations, a slit width of 4 nm will give the best results.

Use a narrow slit (1 or 2 nm) if you want to identify compounds with fine spectral structures or if you need to quantify at high concentrations (> 1000 mAU) with a wavelength at the slope of the spectrum. Signals with a wide bandwidth can be used to reduce baseline noise. Because (digital) bandwidth is computed as average of absorbance, there is no impact on linearity.

Use a wide (8 nm) slit when your sample contains very small concentrations. Always use signals with bandwidth at least as wide as the slit width.

Optimizing Spectral Acquisition

Storage of all spectra consumes a lot of disk space. It is very useful to have all spectra available during optimization of a method or when analyzing unique samples. However when running many samples of the same type, the large size of data files with all spectra may become a burden. The detector provides functions to reduce the amount of data, yet retaining the relevant spectral information.

Range

Only the wavelength range where the compounds in your sample absorb contains information that is useful for purity checks and library searches. Reducing the spectrum storage range saves disk space.

Step

Most substances have broad absorbance bands. Display of spectra, peak purity and library search works best if a spectrum contains 5 to 10 data points per width of the absorbance bands. For anisic acid (the example used before) a step of 4 nm would be sufficient. However a step of 2 nm gives a more optimal display of the spectrum.

Threshold

Sets the peak detector. Only spectra from peaks higher than threshold will be stored when a peak-controlled storage mode is selected.

Margin for Negative Absorbance

The detector adjusts its gain during *balance* such that the baseline may drift slightly negative (about -100 mAU). In some special case, for example, when gradient with absorbing solvents are used, the baseline may drift to more negative values.

Only for such cases, increase the margin for negative absorbance to avoid overflow of the analog-to-digital converter.

Optimizing Selectivity

Quantifying Coeluting Peaks by Peak Suppression

In chromatography, two compounds may often elute together. A conventional dual-signal detector can only detect and quantify both compounds independently from each other if their spectra do not overlap. However, in most cases this is highly unlikely.

With a dual-channel detector based on diode-array technology, quantifying two compounds is possible even when both compounds absorb over the whole wavelength range. The procedure is called peak suppression or signal subtraction. As an example, the analysis of hydrochlorothiazide in the presence of caffeine is described. If hydrochlorothiazide is analyzed in biological samples, there is always a risk that caffeine is present which might interfere chromatographically with hydrochlorothiazide. As the spectra in [Figure 24](#) on page 83 shows, hydrochlorothiazide is best detected at 222 nm, where caffeine also shows significant absorbance. It would therefore be impossible, with a conventional variable wavelength detector, to detect hydrochlorothiazide quantitatively when caffeine is present.




Figure 24 Wavelength Selection for Peak Suppression

With a UV-visible detector based on a diode array and the correct choice of a reference wavelength setting, quantitative detection is possible. To suppress caffeine, the reference wavelength must be set to 282 nm. At this wavelength, caffeine shows exactly the same absorbance as at 222 nm. When the absorbance values are subtracted from each other, any indication of the presence of caffeine is eliminated. In the same way, hydrochlorothiazide can be suppressed if caffeine is to be quantified. In this case the wavelength is set to 204 nm and the reference wavelength to 260 nm. [Figure 25](#) on page 84 shows the chromatographic results of the peak suppression technique.

The trade-off for this procedure is a loss in sensitivity. The sample signal decreases by the absorbance at the reference wavelength relative to the signal wavelength. Sensitivity may be decreased by as much as 10–30 %.




Figure 25 Peak Suppression Using Reference Wavelength

Ratio Qualifiers for Selective Detection of Compound Classes

Ratio qualifiers can be used where, in a complex sample, only one particular class needs to be analyzed – a parent drug and its metabolites in a biological sample, for example. Another example is the selective analysis of derivatives after pre- or post-column derivatization. Specifying a signal ratio that is typical for the sample class is one way of selectively plotting only those peaks that are of interest. The signal output remains at zero so long as the ratio is out of the user-specified ratio range. When the ratio falls within the range, the signal output corresponds to the normal absorbance, giving single, clear peaks on a flat baseline. An example is shown in [Figure 26](#) on page 85 and [Figure 27](#) on page 86.




Figure 26 Wavelength Selection for Ratio Qualifiers




Figure 27 Selectivity by Ratio Qualifiers

In a four-component mixture, only biphenyl was recorded. The other three peaks were suppressed because they did not meet the ratio-qualifier criterion and therefore the output was set to zero. The characteristic wavelengths 249 nm (λ_1) and 224 nm (λ_2) were found from the spectra shown in [Figure 26](#) on page 85. The ratio range was set at 2 – 2.4 (2.2 $\pm$ 10%). Only when the ratio between 249 and 224 nm was within this range, is the signal plotted. Of all four peaks, only the third fulfilled the criterion ([Figure 27](#) on page 86). The others were not plotted.

Optimizing the Detector Regarding to the System

Delay Volume and Extra-Column Volume

The *delay volume* is defined as the system volume between the point of mixing in the pump and the top of the column.

The *extra-column volume* is defined as the volume between the injection point and the detection point, excluding the volume in the column.

Extra-Column Volume

Extra-column volume is a source of peak dispersion that will reduce the resolution of the separation and so should be minimized. Smaller diameter columns require proportionally smaller extra-column volumes to keep peak dispersion at a minimum.

In a liquid chromatograph the extra-column volume will depend on the connection tubing between the autosampler, column and detector; and on the volume of the flow cell in the detector. The extra-column volume is minimized with the Agilent 1290 Infinity/Agilent 1260 Infinity LC system due to the narrow-bore (0.12 mm i.d.) tubing, the low-volume heat exchangers in the column compartment and the Max-Light cartridge cell in the detector.

How to Configure the Optimum Delay Volume

To maintain resolution in the Agilent 1290 Infinity II DAD/Agilent 1290 Infinity DAD FS the 10 mm Max-Light cartridge cell has a low dispersion volume (σ volume 1.0 μL) and no further volume optimization is required. In situations where the alternative 60 mm Max-Light high sensitivity cell is used to get higher sensitivity the cell volume is optimized for the use with 3 mm and 4.6 mm inner diameter columns.

How to Achieve Higher Sensitivity

The detector has a number of parameters that are used to optimize performance. The following sections describe how the detector parameters affect performance characteristics:

- Flow cell affects sensitivity,
- Wavelength and bandwidth affect sensitivity, selectivity and linearity,
- Slit width affects sensitivity, spectral resolution and linearity,
- Peak width affects sensitivity and resolution.

Flow Cell

The Max-Light cartridge flow cell has a standard 10 mm path length and is optimized for minimal volume and dispersion (σ volume 1.0 μL). It has high light transmission minimizing noise to reduce noise due to the optofluidic waveguide. It is suitable for use with a wide range of analytical columns from short narrow-bore columns to long standard diameter (4.6 mm) columns. Generally the peak dispersion volume (calculated from peak width x flow rate) should be greater than about 2 μL for this cell (for example 0.02 min x 200 $\mu\text{L}/\text{min}$ = 4 μL).

The Max-Light high sensitivity cell has a path length of 60 mm and this will give between three and five times increase in signal-to-noise values depending on the application conditions. The dispersion volume is fractionally increased compared to the standard cell.

Wavelength and Bandwidth

The detector measures absorbance simultaneously at wavelengths from 190 nm to 640 nm using diode-array detection. A UV-lamp provides good sensitivity over the whole wavelength range. The diode-array detector (DAD) can simultaneously compute and send to the data system up to eight chromatographic signals and the full-range spectra at every time point.

A UV chromatogram or signal is a plot of absorbance data versus time and is defined by its wavelength and bandwidth.

- The wavelength indicates the center of the detection band.
- The bandwidth defines the wavelength range over which the absorbance values are averaged to give the result at each time point.

For example, a signal at wavelength 250 nm with a bandwidth of 16 nm will be an average of the absorbance data from 242 nm to 258 nm. Additionally, a reference wavelength and reference bandwidth can be defined for each signal. The average absorbance from the reference bandwidth centered on the reference wavelength will be subtracted from its equivalent value at the signal wavelength to produce the output chromatogram.

The signal wavelength and bandwidth can be chosen so that they are optimized for:

- Broad band universal detection
- Narrow band selective detection
- Sensitivity for a specific analyte.

Broad band or universal detection works by having a wide bandwidth to detect any species with absorbance in that range. For example, to detect all absorbing molecules between 200 nm and 300 nm set a signal at 250 nm with a bandwidth of 100 nm. The disadvantage is that sensitivity will not be optimal for any one of those molecules. Narrow band or selective detection is used most often. The UV spectrum for a particular molecule is examined and an appropriate absorbance maximum is selected. If possible, the range where solvents absorb strongly should be avoided (below 220 nm for methanol, below 210 nm for acetonitrile). For example, in [Figure 28](#) on page 91, anisic acid has a suitable absorbance maximum at 252 nm. A narrow bandwidth of 4 nm to 12 nm generally gives good sensitivity and is specific for absorbance in a narrow range.

The narrow band can be optimized for sensitivity for a specific molecule. As the bandwidth is increased the signal is reduced but so is the noise and there will be an optimum for best S/N. As an approximate guide, this optimum is often close to the natural bandwidth at half-height of the absorption band in the UV spectrum. In the anisic acid example this is 30 nm.

The analytical wavelength is usually set at a wavelength maximum to increase sensitivity to that molecule. The detector is linear up to 2 AU and beyond for many applications. This offers a wide linear range for concentration. For high concentration analysis the concentration linear range can be extended by setting the wavelength to one with a lower absorbance such as a wavelength minimum or by taking a wider bandwidth which usually includes lower absorbance values. The use of wavelength maxima and minima for quantitation dates back to

conventional UV detectors which because of mechanical tolerances in moving gratings needed to avoid steeply sloping parts of the spectrum. Diode-array based detectors do not have this limitation but for reasons of convention maxima and minima are chosen in preference to other parts of the spectrum.

The reference bandwidth is normally set on a region of the UV spectrum in which the analyte has no absorbance. This is shown in the spectrum for anisic acid in [Figure 28](#) on page 91. This spectrum is typical of many small molecules containing a UV chromophore. For best results the reference has been set so that it is a wide band as close to the signal wavelength as possible but on a zero absorbance region. Reference bandwidths of 60 nm to 100 nm are commonly used. The default reference is 360 nm with a bandwidth of 100 nm. A wide bandwidth is used because this reduces the noise in the reference signal (from statistical theory, the error, i.e. noise in this case, is reduced by the square root of the number of determinations). It is important that the reference bandwidth does not extend to a part of the spectrum that has some absorbance as this would then reduce the resulting signal and sensitivity would be reduced. The use of a reference wavelength can help to reduce drift or wander in the chromatogram caused by refractive index changes due to room temperature fluctuation or gradient operation. The effect of a reference signal can be easily tested by setting two otherwise identical signals, one with and one without a reference signal. If there is no part of the spectrum with zero absorbance then it will be better to have the reference signal turned off.




Figure 28 Spectrum of Anisic Acid

Peak Width, Response Time and Data Collection Rate

The peak width setting, response time and data rate in the detector are all linked. The available settings are shown in [Table 12](#) on page 92 and in [Table 13](#) on page 93. It is important to set this correctly for optimum sensitivity and to preserve the resolution achieved in the separation.

The detector internally acquires data points faster than is needed for a chromatogram and processes them to produce the signal seen by the data system. Part of the processing reduces the data to an appropriate data rate which allows the chromatographic peaks to be accurately drawn. As with most analytical determinations groups of readings are effectively averaged to reduce error in the result. The detector bunches raw data points and produces the output signal data at the required data collection rate by an electronic filtering process. If the resulting data rate is too slow (over filtering) the peak heights will be reduced and the resolution between them reduced; too fast and the data is noisier than it need be to accurately profile narrow peaks.

The *peak width* setting in the detector allows the user to correctly set these parameters without needing any knowledge other than sight of the chromatogram integration results to see how wide the peaks are. The peak width setting should be set for the narrowest peak width observed in the

5 Optimizing the Detector

Optimizing the Detector Regarding to the System

chromatogram. If it is set too wide it will make the peaks appear lower in height and wider (and potentially less resolved) and if it is set too narrow it will increase the baseline noise unnecessarily. Essentially the software uses this value to set the *data collection rate* such that it collects enough data points over the narrowest peaks and it is aiming for 15 to 25 points across a peak. The 1290 Infinity DAD can collect at a maximum up to 240 Hz if required which would allow enough data points to be collected over a peak that is only 0.1 s wide. The *response time* setting is another way of indicating how this filtering is set. It is measured in seconds and is about one-third of the peak width value (which is measured in minutes). It effectively shows how quickly the plotted signal responds to a step change in the input signal.

NOTE

The full spectra is not available under all conditions.

Based on the data points, the scan data rate is reduced, see [Table 12](#) on page 92 and [Table 13](#) on page 93.

Table 12 Peak Width — Response Time — Data Rate (G7117B)

Peak width at half height [min] ¹	Response [s]	Signal data rate [Hz]	Scan data rate [HZ] ≤126 pts/scan	Scan data rate [HZ] ≤251 pts/scan	Scan data rate [HZ] ≤501 pts/scan	Scan data rate [HZ] >501 pts/scan
< 0.00078125	0.0078125	240	240	80	40	20
> 0.00078125	0.015625	240	240	80	40	20
> 0.0015625	0.03125	160	160	80	40	20
> 0.003125	0.0625	80	80	80	40	20
> 0.00625	0.125	40	40	40	40	20
> 0.0125	0.25	20	20	20	20	20
> 0.025	0.5	10	10	10	10	10
> 0.05	1	5	5	5	5	5
> 0.1	2	2.5	2.5	2.5	2.5	2.5
> 0.2	4	1.25	1.25	1.25	1.25	1.25
> 0.4	8	0.625	0.625	0.625	0.625	0.625
> 0.85	16	0.3125	0.3125	0.3125	0.3125	0.3125

¹ Values in the User Interface may be rounded.

Table 13 Peak Width — Response Time — Data Rate (G7117A)

	Peak width at half height [min] ¹	Response [s]	Scan data rate[Hz] ≤251 pts/scan	Scan data rate[Hz] ≤501 pts/scan	Scan data rate[Hz] > 501 pts/scan
< 0.0015625	0.015625	120	120	40	20
> 0.0015625	0.03125	120	120	40	20
> 0.003125	0.0625	80	80	40	20
> 0.00625	0.125	40	40	40	20
> 0.0125	0.25	20	20	20	20
> 0.025	0.5	10	10	10	10
> 0.05	1	5	5	5	5
> 0.1	2	2.5	2.5	2.5	2.5
> 0.2	4	1.25	1.25	1.25	1.25
> 0.4	8	0.625	0.625	0.625	0.625
> 0.85	16	0.3125	0.3125	0.3125	0.3125

¹ Values in the User Interface may be rounded.

NOTE

The maximum spectra scan rate depends on the data points per scan, see [Table 12](#) on page 92 and [Table 13](#) on page 93. Running at 240, the spectra scan data rate is reduced automatically if the spectra scan data rate is more than 251 points/scan.

Warm up of the Detector

Give the optical unit enough time to warm-up and stabilize (> 60 minutes). The detector is temperature controlled. After turn-on of the detector, it goes through a cycle of different states:

- 0 to 0.5 minutes the heater control is OFF and the heater element runs at 0 % duty cycle.
- 0.5 to 1 minutes the heater control is OFF and the heater element runs at 66% duty cycle. This first minute is used as self-test of the heater functionality.
- 1 to 30 minutes the heater control is OFF and the heater element runs at 40% duty cycle.
- After 30 minutes the heater control is ON and is working with optimized parameters to get the optical unit into the optimal temperature window stabilized.

This cycle starts

- when the detector is turned off/on
- when the lamp is turned off/on

to ensure that the temperature control operates in a defined control range.

NOTE

The times to stabilize the baseline may vary from instrument to instrument and depends on the environment. The example below was done under stable environmental conditions.

The figures below show the first two hours of a detector warm-up phase. The lamp was turned on immediately after turn on of the detector.




Figure 29 Detector Warm-up – 1st hour





Figure 30 Detector Warm-up – 2nd hour

5 Optimizing the Detector

Warm up of the Detector



6 Troubleshooting and Diagnostics

Available Tests vs User Interfaces 98

Agilent Lab Advisor Software 99

Overview about the troubleshooting and diagnostic features.



Available Tests vs User Interfaces


NOTE

Depending on the used interface, the available tests and the screens/reports may vary. Preferred tool should be the Agilent Lab Advisor, see “[Agilent Lab Advisor Software](#)” on page 99.

Agilent Lab Advisor B.02.06 or later is required.

The Instant Pilot does not support the G7117A/B at introduction. If running as G4212A/B DAD (emulation mode) the Instant Pilot firmware must be B.02.16.

- Preferred tool should be the Agilent Lab Advisor software, see “[Agilent Lab Advisor Software](#)” on page 99.
- Screenshots used within these procedures are based on the Agilent Lab Advisor software.



Agilent Lab Advisor Software

The Agilent Lab Advisor Software is a standalone product that can be used with or without chromatographic data system. Agilent Lab Advisor helps to manage the lab for high-quality chromatographic results by providing a detailed system overview of all connected analytical instruments with instrument status, Early Maintenance Feedback counters (EMF), instrument configuration information, and diagnostic tests. By the push of a button, a detailed diagnostic report can be generated. Upon request, the user can send this report to Agilent for a significantly improved troubleshooting and repair process.

The Agilent Lab Advisor software is available in two versions:


- Lab Advisor Basic
- Lab Advisor Advanced

Lab Advisor Basic is included with every Agilent 1200 Infinity Series and Infinity II Series pump.

The Lab Advisor Advanced features can be unlocked by purchasing a license key, and include real-time monitoring of instrument actuals, all various instrument signals, and state machines. In addition, all diagnostic test results, calibration results, and acquired signal data can be uploaded to a shared network folder. The Review Client included in Lab Advisor Advanced allows to load and examine the uploaded data no matter on which instrument it was generated. This makes Data Sharing an ideal tool for internal support groups and users who want to track the instrument history of their analytical systems.

The optional Agilent Maintenance Wizard Add-on provides an easy-to-use, step-by-step multimedia guide for performing preventive maintenance on Agilent 1200 Infinity and Infinity II Series instruments.

The tests and diagnostic features that are provided by the Agilent Lab Advisor software may differ from the descriptions in this manual. For details, refer to the Agilent Lab Advisor software help files.



7 Error Information

What Are Error Messages	103
General Error Messages	104
Timeout	104
Shutdown	104
Remote Timeout	105
Lost CAN Partner	106
Leak Sensor Short	106
Leak Sensor Open	107
Compensation Sensor Open	107
Compensation Sensor Short	108
Fan Failed	108
Leak	109
Open Cover	109
Cover Violation	110
ERI Messages	110
Detector Error Messages	111
Diode Current Leakage	111
UV Lamp Current	111
UV Lamp Voltage	112
UV Ignition Failed	112
UV Heater Current	113
Calibration Values Invalid	113
Wavelength Recalibration Lost	114
Illegal Temperature Value from Sensor on Main Board	114
Illegal Temperature Value from Sensor at Air Inlet	115



Heater at fan assembly failed [115](#)

Heater Power At Limit [116](#)

This chapter describes the meaning of error messages, and provides information on probable causes and suggested actions how to recover from error conditions.

What Are Error Messages

Error messages are displayed in the user interface when an electronic, mechanical, or hydraulic (flow path) failure occurs which requires attention before the analysis can be continued (for example, repair, or exchange of consumables is necessary). In the event of such a failure, the red status indicator at the front of the module is switched on, and an entry is written into the module logbook.

If an error occurs outside a method run, other modules will not be informed about this error. If it occurs within a method run, all connected modules will get a notification, all LEDs get red and the run will be stopped. Depending on the module type, this stop is implemented differently. For example, for a pump the flow will be stopped for safety reasons. For a detector, the lamp will stay on in order to avoid equilibration time. Depending on the error type, the next run can only be started, if the error has been resolved, for example liquid from a leak has been dried. Errors for presumably single time events can be recovered by switching on the system in the user interface.

Special handling is done in case of a leak. As a leak is a potential safety issue and may have occurred at a different module from where it has been observed, a leak always causes a shutdown of all modules, even outside a method run.

In all cases, error propagation is done via the CAN bus or via an APG remote cable (see documentation for the APG interface).

General Error Messages

General error messages are generic to all Agilent series HPLC modules and may show up on other modules as well.

Timeout

Error ID: 0062

The timeout threshold was exceeded.

Probable cause

- 1** The analysis was completed successfully, and the timeout function switched off the module as requested.
- 2** A not-ready condition was present during a sequence or multiple-injection run for a period longer than the timeout threshold.

Suggested actions

Check the logbook for the occurrence and source of a not-ready condition. Restart the analysis where required.

Check the logbook for the occurrence and source of a not-ready condition. Restart the analysis where required.

Shutdown

Error ID: 0063

An external instrument has generated a shutdown signal on the remote line.

The module continually monitors the remote input connectors for status signals. A LOW signal input on pin 4 of the remote connector generates the error message.

Probable cause	Suggested actions
1 Leak detected in another module with a CAN connection to the system.	Fix the leak in the external instrument before restarting the module.
2 Leak detected in an external instrument with a remote connection to the system.	Fix the leak in the external instrument before restarting the module.
3 Shut-down in an external instrument with a remote connection to the system.	Check external instruments for a shut-down condition.
4 The degasser failed to generate sufficient vacuum for solvent degassing.	Check the vacuum degasser for an error condition. Refer to the <i>Service Manual</i> for the degasser or the 1260 pump that has the degasser built-in.

Remote Timeout

Error ID: 0070

A not-ready condition is still present on the remote input. When an analysis is started, the system expects all not-ready conditions (for example, a not-ready condition during detector balance) to switch to run conditions within one minute of starting the analysis. If a not-ready condition is still present on the remote line after one minute the error message is generated.

Probable cause	Suggested actions
1 Not-ready condition in one of the instruments connected to the remote line.	Ensure the instrument showing the not-ready condition is installed correctly, and is set up correctly for analysis.
2 Defective remote cable.	Exchange the remote cable.
3 Defective components in the instrument showing the not-ready condition.	Check the instrument for defects (refer to the instrument's documentation).

Lost CAN Partner

Error ID: 0071

During an analysis, the internal synchronization or communication between one or more of the modules in the system has failed.

The system processors continually monitor the system configuration. If one or more of the modules is no longer recognized as being connected to the system, the error message is generated.

Probable cause

- 1 CAN cable disconnected.
- 2 Defective CAN cable.
- 3 Defective main board in another module.

Suggested actions

- Ensure all the CAN cables are connected correctly.
 - Ensure all CAN cables are installed correctly.
- Exchange the CAN cable.
- Switch off the system. Restart the system, and determine which module or modules are not recognized by the system.

Leak Sensor Short

Error ID: 0082

The leak sensor in the module has failed (short circuit).

The current through the leak sensor is dependent on temperature. A leak is detected when solvent cools the leak sensor, causing the leak sensor current to change within defined limits. If the current increases above the upper limit, the error message is generated.

Probable cause

- 1 Defective leak sensor.
- 2 Leak sensor incorrectly routed, being pinched by a metal component.

Suggested actions

- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Leak Sensor Open

Error ID: 0083

The leak sensor in the module has failed (open circuit).

The current through the leak sensor is dependent on temperature. A leak is detected when solvent cools the leak sensor, causing the leak-sensor current to change within defined limits. If the current falls outside the lower limit, the error message is generated.

Probable cause

- 1 Leak sensor not connected to the Power Switch board.
- 2 Defective leak sensor.
- 3 Leak sensor incorrectly routed, being pinched by a metal component.

Suggested actions

- Please contact your Agilent service representative.
- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Compensation Sensor Open

Error ID: 0081

The ambient-compensation sensor (NTC) on the power switch board in the module has failed (open circuit).

The resistance across the temperature compensation sensor (NTC) on the power switch board is dependent on ambient temperature. The change in resistance is used by the leak circuit to compensate for ambient temperature changes. If the resistance across the sensor increases above the upper limit, the error message is generated.

Probable cause

- 1 Loose connection between the power switch board and the main board
- 2 Defective power switch board

Suggested actions

- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Compensation Sensor Short

Error ID: 0080

The ambient-compensation sensor (NTC) on the power switch board in the module has failed (open circuit).

The resistance across the temperature compensation sensor (NTC) on the power switch board is dependent on ambient temperature. The change in resistance is used by the leak circuit to compensate for ambient temperature changes. If the resistance across the sensor falls below the lower limit, the error message is generated.

Probable cause	Suggested actions
1 Defective power switch board	Please contact your Agilent service representative.
2 Loose connection between the power switch board and the main board	Please contact your Agilent service representative.

Fan Failed

Error ID: 0068

The cooling fan in the module has failed.

The hall sensor on the fan shaft is used by the main board to monitor the fan speed. If the fan speed falls below a certain limit for a certain length of time, the error message is generated.

Depending on the module, assemblies (e.g. the lamp in the detector) are turned off to assure that the module does not overheat inside.

Probable cause	Suggested actions
1 Fan cable disconnected.	Please contact your Agilent service representative.
2 Defective fan.	Please contact your Agilent service representative.
3 Defective main board.	Please contact your Agilent service representative.

Leak

Error ID: 0064

A leak was detected in the module.

The signals from the two temperature sensors (leak sensor and board-mounted temperature-compensation sensor) are used by the leak algorithm to determine whether a leak is present. When a leak occurs, the leak sensor is cooled by the solvent. This changes the resistance of the leak sensor which is sensed by the leak-sensor circuit on the main board.

Probable cause

- 1 Loose fittings.
- 2 Broken capillary.

Suggested actions

- Ensure all fittings are tight.
- Exchange defective capillaries.

Open Cover

Error ID: 0205

The top foam has been removed.

Probable cause

- 1 Foam not activating the sensor.
- 2 Defective sensor or main board.

Suggested actions

- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Cover Violation

Error ID: 7461

The top foam has been removed.

The sensor on the main board detects when the top foam is in place. If the foam is removed while the lamps are on (or if an attempt is made to switch on for example the lamps with the foam removed), the lamps are switched off, and the error message is generated.

Probable cause

- 1 The top foam was removed during operation.
- 2 Foam not activating the sensor.

Suggested actions

Please contact your Agilent service representative.

Please contact your Agilent service representative.

ERI Messages

Error ID: 11120 (+5 V) , 11121 (+25 V)

The ERI (Enhanced Remote Interface) provides two error events related to over current situations on the +5 V and +24 V lines.

Probable cause

- 1 The load on the ERI is too high.

Suggested actions

Reduce the load.

Detector Error Messages

These errors are detector specific.

Diode Current Leakage

Error ID: 1041

When the detector is switched on, the processor checks the leakage current of each of the optical diodes. If the leakage current exceeds the upper limit, the error message is generated.

Probable cause	Suggested actions
1 Defective PDA/optical unit.	Please contact your Agilent service representative.
2 Defective connector or cable.	Please contact your Agilent service representative.

UV Lamp Current

Error ID: 7450

The UV lamp current is missing.

The processor continually monitors the anode current drawn by the lamp during operation. If the anode current falls below the lower current limit, the error message is generated.

Probable cause	Suggested actions
1 Lamp disconnected.	Ensure the UV lamp connector is seated firmly.
2 Defective UV lamp or non-Agilent lamp.	Exchange the UV lamp.
3 Defective detector main board.	Please contact your Agilent service representative.
4 Defective power supply.	Please contact your Agilent service representative.

UV Lamp Voltage

Error ID: 7451

The UV lamp anode voltage is missing.

The processor continually monitors the anode voltage across the lamp during operation. If the anode voltage falls below the lower limit, the error message is generated.

Probable cause	Suggested actions
1 Defective UV lamp or non-Agilent lamp.	Exchange the UV lamp.
2 Defective detector main board.	Please contact your Agilent service representative.
3 Defective power supply.	Please contact your Agilent service representative.

UV Ignition Failed

Error ID: 7452

The UV lamp failed to ignite.

The processor monitors the UV lamp current during the ignition cycle. If the lamp current does not rise above the lower limit within 2 – 5 seconds, the error message is generated.

Probable cause	Suggested actions
1 Lamp too hot. Hot gas discharge lamps may not ignite as easily as cold lamps.	Switch off the lamp and allow it to cool down for at least 15 minutes.
2 Lamp disconnected.	Ensure the lamp is connected.
3 Defective UV lamp or non-Agilent lamp.	Exchange the UV lamp.
4 Defective detector main board.	Please contact your Agilent service representative.
5 Defective power supply.	Please contact your Agilent service representative.

UV Heater Current

Error ID: 7453

The UV lamp heater current is missing.

During UV lamp ignition, the processor monitors the heater current. If the current does not rise above the lower limit within one second, the error message is generated.

Probable cause	Suggested actions
1 Lamp disconnected.	Ensure the UV lamp is connected.
2 Ignition started without the top foam in place.	Please contact your Agilent service representative.
3 Defective UV lamp or non-Agilent lamp.	Exchange the UV lamp.
4 Defective detector main board.	Please contact your Agilent service representative.
5 Defective power supply.	Please contact your Agilent service representative.

Calibration Values Invalid

Error ID: 1036

The calibration values read from the spectrometer ROM are invalid.

After recalibration, the calibration values are stored in ROM. The processor periodically checks if the calibration data are valid. If the data are invalid or cannot be read from the spectrometer ROM, the error message is generated.

Probable cause	Suggested actions
1 Defective connector or cable.	Please contact your Agilent service representative.
2 Defective PDA/optical unit.	Please contact your Agilent service representative.

Wavelength Recalibration Lost

Error ID: 1037

The calibration information needed for your detector to operate correctly has been lost.

During calibration of the detector the calibration values are stored in ROM. If no data is available in the spectrometer ROM, the error message is generated.

Probable cause

- 1 The detector is new.
- 2 The detector has been repaired.

Suggested actions

Recalibrate the detector.

Please contact your Agilent service representative.

Illegal Temperature Value from Sensor on Main Board

Error ID: 1071

This temperature sensor (located on the detector main board) delivered a value outside the allowed range. The parameter of this event equals the measured temperature in 1/100 centigrade. As a result the temperature control is switched off.

Probable cause

- 1 Defective sensor or main board.
- 2 Detector is exposed to illegal ambient conditions.

Suggested actions

Please contact your Agilent service representative.

Verify that the ambient conditions are within the allowed range.

Illegal Temperature Value from Sensor at Air Inlet

Error ID: 1072

This temperature sensor delivered a value outside the allowed range. The parameter of this event equals the measured temperature in 1/100 centigrade. As a result the temperature control is switched off.

Probable cause

- 1 The temperature sensor is defect.
- 2 Detector is exposed to illegal ambient conditions.

Suggested actions

- Replace the cable to the main board.
 - Please contact your Agilent service representative.
- Verify that the ambient conditions are within the allowed range.

Heater at fan assembly failed

Error ID: 1073

Every time the deuterium lamp or the tungsten lamp (DAD only) is switched on or off a heater self-test is performed. If the test fails an error event is created. As a result the temperature control is switched off.

Probable cause

- 1 Defective connector or cable.
- 2 Defective heater.

Suggested actions

- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Heater Power At Limit

Error ID: 1074

The available power of the heater reached either the upper or lower limit. This event is sent only once per run. The parameter determines which limit has been hit:

0 means upper power limit hit (excessive ambient temperature drop).


1 means lower power limit hit (excessive ambient temperature increase).

Probable cause

- 1 Excessive ambient temperature change.

Suggested actions

Wait until temperature control equilibrates.



8 Test Functions and Calibration

Introduction	118
Use of Max-Light Cartridge Test Cell	120
Conditions of Detector	121
Failing a Test	121
Self-Test	122
Intensity Test	124
Test Failed	126
Cell Test	127
Test Failed (low ratio value)	129
Quick Noise Test	130
Test Failed	132
ASTM Drift and Noise Test	133
Test Failed	135
Slit Test	136
Slit Test (G7117B)	136
Test Failed	138
Slit Test (G7117A)	138
Wavelength Verification Test	139
Wavelength Calibration	141
Wavelength Recalibration Fails	143
D/A Converter (DAC) Test	144
Test Evaluation	146
Dark Current Test	147
Test Failed	148
Other Lab Advisor Functions	149
EMFs - Early Maintenance Feature	149

This chapter describes the tests for the module.



Introduction

All tests are described based on the Agilent Lab Advisor Software B.02.06. Other user interfaces may not provide any test or just a few.

Table 14 Interfaces and available test functions

Interface	Comment	Available Function
Agilent Instrument Utilities	Maintenance tests are available	• Intensity • Cell • WL Calibration
Agilent Lab Advisor	All tests are available	• Self-Test • Intensity • Quick Noise • ASTM Drift and Noise • Cell • Dark Current • D/A Converter • Slit (G7117B only) • WL Verification • WL Calibration • Test Chromatogram (Tools) • Spectra Scan (Tools) • Module Infos (Tools) • Diagnostic (Tools)
Agilent ChemStation	No tests available Adding of temperature/lamp signals to chromatographic signals possible	• Temperature main board • Temperature optical unit • Lamp anode voltage

For details on the use of the interface refer to the interface documentation.

The Lab Advisor shows the available test under Service & Diagnostics.




Figure 31 The Lab Advisor shows the available test

Use of Max-Light Cartridge Test Cell

The Max-Light Cartridge Test Cell is recommended to be used for several tests instead of the Max-Light Cartridge Cell (10 mm, $V(\sigma) = 1 \mu\text{L}$) or the Max-Light Cartridge Cell (60 mm, $V(\sigma) = 4 \mu\text{L}$) because it allows running the test(s) without any influence of the rest of the system (degasser, pump, sampler and others).

The results of the test cell are comparable with the Max-Light Cartridge Cell (10 mm, $V(\sigma) = 1 \mu\text{L}$) filled with water, e.g. Intensity Profile. Only the Absorbance value is higher on the Max-Light Cartridge Cell.

If the profile of the Max-Light Cartridge Cell differs in the low UV range, then absorbing solvents are in the cell and should be flushed out. See also “Clean the Max-Light Cartridge Cell” on page 163.

NOTE

When using the Max-Light Cartridge Cell for tests/calibrations, it should be run at 0.5 mL/min constant flow with water. This assures that the light path is always flushed.

Below table gives an idea on the signal height variation of the Max-Light Cartridge Cells compared to Max-Light Cartridge Test Cell.

Table 15 Max-Light Cartridge Cells compared to Max-Light Cartridge Test Cell

Part Number	Description	Signal Height (typical)
G4212-60011	Max-Light Cartridge Test Cell	100 %
G4212-60008	Max-Light Cartridge Cell 10 mm $V(\sigma) = 1 \mu\text{L}$	~ 100 %
G4212-60007	Max-Light Cartridge Cell 60 mm $V(\sigma) = 4 \mu\text{L}$	~ 100 %
G4212-60032	Max-Light Cartridge Cell HDR (3.7 mm, $V(\sigma) 0.4 \mu\text{L}$)	100 %
G4212-60017	Max-Light Cartridge Cell ULD (10 mm, $V(\sigma) 0.6 \mu\text{L}$)	100 %

Conditions of Detector

The test usually should be performed with a detector turned on for at least one hour, so that the temperature regulation of the optical unit is working (not active during the first 30 minutes after turn on). If the detector is on, tests can be performed usually 10 minutes after the UV-lamp has been turned on.

Failing a Test

If a test fails with the Max-Light Cartridge Cell repeat the test with the Max-Light Cartridge Test Cell and compare. If the test fails also, then start with proposed actions mentioned in the details of the tests.

Self-Test

The self-test runs a series of individual tests (described on the next pages), and evaluates the results automatically. The following tests are run:

- Slit Test (G7117B only)
- Dark Current Test
- Intensity Test
- Wavelength Verification Test
- ASTM Noise Test, a simplified version of the ASTM Drift and Noise Test (without testing the Drift)

When For complete detector check.

Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water)
OR	1	Max-Light Cartridge Test Cell

- Preparations**
- Lamp must be on for at least 10 minutes.
 - For noise test a longer warm-up time may be required (> 2 hours).
 - When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.

- 1 Run the **Self-Test** with Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 32 Self-Test – Results

Under the tab **Signals** you can find the detailed signals from the tests.

Intensity Test

The intensity test measures the intensity of the UV-lamp over the full wavelength range (190 - 640 nm). Four spectral ranges are used to evaluate the intensity spectrum. The test is used to determine the performance of the lamp and optics (see also “Cell Test” on page 127). When the test is started, the 1-nm slit is moved into the light path automatically (G7117B only). On the G7117A, the 4 nm fixed slit is used. To eliminate effects due to absorbing solvents, the test should be done with water in the Max-Light Cartridge Cell or with the Max-Light Cartridge Test Cell. The shape of the intensity spectrum is primarily dependent on the lamp, grating, and diode array characteristics. Therefore, intensity spectra will differ slightly between instruments.

When	In case of UV-lamp problem (drift, noise).	
Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water)
	OR	1
Preparations	1	Max-Light Cartridge Test Cell
	Lamp must be on for at least 10 minutes.	

- 1 Run the **Intensity-Test** with Agilent Lab Advisor (for further information see Online-Help of user interface).

Test Name	Intensity Test	Description	The test scans the Intensity spectrum generated by the UV Lamp.																												
Module	G4212A:PR00100015																														
Status	Passed																														
Start Time	7/9/2009 2:14:09 PM																														
Stop Time	7/9/2009 2:14:30 PM																														
																															
Test Procedure		Result																													
 1. Check Prerequisites...																															
 2. Insert supported Cell or Test Cell.																															
 3. Scan Intensity Spectrum...																															
 4. Evaluate Data...																															
		<table><tr><th>Name</th><th>Value</th></tr><tr><td>Cell Product Number</td><td>G4212-60011</td></tr><tr><td>Cell Name</td><td>Max-Light Test Cell</td></tr><tr><td>Cell Type</td><td>10 mm/0 µl</td></tr><tr><td>Lamp Type</td><td>Automatic Mode</td></tr><tr><td>Lowest Intensity in Range 190 - 220 nm</td><td>30261 Counts</td></tr><tr><td>Lowest Intensity in Range 221 - 350 nm</td><td>35197 Counts</td></tr><tr><td>Lowest Intensity in Range 351 - 500 nm</td><td>8211 Counts</td></tr><tr><td>Lowest Intensity in Range 501 - 640 nm</td><td>2201 Counts</td></tr><tr><td>Highest Intensity in Range 190 - 350 nm</td><td>151611 Counts</td></tr><tr><td>Highest Intensity in Range 351 - 500 nm</td><td>38272 Counts</td></tr><tr><td>Highest Intensity in Range 501 - 640 nm</td><td>38691 Counts</td></tr><tr><td>Spectrum Integral</td><td>16465136</td></tr><tr><td>UV Integral (190 - 349 nm)</td><td>12108449</td></tr></table>		Name	Value	Cell Product Number	G4212-60011	Cell Name	Max-Light Test Cell	Cell Type	10 mm/0 µl	Lamp Type	Automatic Mode	Lowest Intensity in Range 190 - 220 nm	30261 Counts	Lowest Intensity in Range 221 - 350 nm	35197 Counts	Lowest Intensity in Range 351 - 500 nm	8211 Counts	Lowest Intensity in Range 501 - 640 nm	2201 Counts	Highest Intensity in Range 190 - 350 nm	151611 Counts	Highest Intensity in Range 351 - 500 nm	38272 Counts	Highest Intensity in Range 501 - 640 nm	38691 Counts	Spectrum Integral	16465136	UV Integral (190 - 349 nm)	12108449
Name	Value																														
Cell Product Number	G4212-60011																														
Cell Name	Max-Light Test Cell																														
Cell Type	10 mm/0 µl																														
Lamp Type	Automatic Mode																														
Lowest Intensity in Range 190 - 220 nm	30261 Counts																														
Lowest Intensity in Range 221 - 350 nm	35197 Counts																														
Lowest Intensity in Range 351 - 500 nm	8211 Counts																														
Lowest Intensity in Range 501 - 640 nm	2201 Counts																														
Highest Intensity in Range 190 - 350 nm	151611 Counts																														
Highest Intensity in Range 351 - 500 nm	38272 Counts																														
Highest Intensity in Range 501 - 640 nm	38691 Counts																														
Spectrum Integral	16465136																														
UV Integral (190 - 349 nm)	12108449																														

Figure 33 Intensity Test – Results




Figure 34 Intensity Test – Signals

Test Failed

Intensity Test Evaluation

Probable cause

- 1** Absorbing solvent or air bubble in flow cell.
- 2** Incorrect calibration
- 3** Dirty or contaminated flow cell.
- 4** Dirty or contaminated optical components.
- 5** Old UV-lamp.
- 6** Defect optical unit.

Suggested actions

- Ensure the flow cell is filled with water, and free from air bubbles.
 - Repeat test with Max-Light Cartridge Test Cell and compare results.
- Recalibrate and repeat the test.
- Run the cell test. If the test fails, flush the flow cell. See also [“Clean the Max-Light Cartridge Cell”](#) on page 163.
- Please contact your Agilent service representative.
- Exchange the UV-lamp.
- If the test fails with Max-Light Cartridge Test Cell and new UV-lamp, please contact your Agilent service representative.

NOTE

If only one range fails and the application does not require this range, the lamp may not be changed.

Cell Test

The cell test measures the intensity of the UV-lamp over the full wavelength range (190 - 690 nm), once with the Max-Light Cartridge Cell installed, and once with the Max-Light Cartridge Test Cell. The resulting intensity ratio is a measure of the amount of light absorbed by the Max-Light Cartridge flow cell. The test can be used to check for dirty or contaminated flow cell windows. When the test is started, the 1-nm slit is moved into the light path automatically (G7117B only). On the G7117A, the 4 nm fixed slit is used.

This test should be performed initially with a new detector/flow cell. The values should be kept for later reference/comparison.

When In case of low intensity or noise and drift problem.

Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water)
	1	Max-Light Cartridge Test Cell

- Preparations**
- Lamp must be on for at least 10 minutes.
 - When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.

8 Test Functions and Calibration

Cell Test

- 1 Run the **Cell-Test** with Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 35 Cell Test – Results




Figure 36 Cell Test – Signals (example shows low intensity for flow cell)

Test Failed (low ratio value)

Cell Test Evaluation

Probable cause

- 1 Absorbing solvent or air bubble in flow cell.
- 2 Dirty or contaminated flow cell.

Suggested actions

Ensure the flow cell is filled with water, and free from air bubbles.

Clean the flow cell as described in [“Clean the Max-Light Cartridge Cell”](#) on page 163.

Quick Noise Test

The quick noise test measures the noise of the detector, with Max-Light Cartridge Cell or with Max-Light Cartridge Test Cell installed, in one minute intervals over a total of 5 minutes.

The noise of the detector is calculated by using the maximum amplitude for all random variations of the detector signal of frequencies greater than one cycle per hour. The noise is determined for 5 one minute intervals and is based on the accumulated peak-to-peak noise for the intervals. At least seven data points per cycles are used in the calculation. The cycles in the noise determination are not overlapping.

If the test is performed with the Max-Light Cartridge Test Cell, the test results are not influenced by solvent or pump effects.

When	In case of noise and drift problem.	
Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water)
	OR	1 Max-Light Cartridge Test Cell
Preparations	• Detector and UV-lamp must be on for at least 2 hours. • ASTM measurements based on specifications may require longer stabilization times. • When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.	

- 1 Run the **Quick Noise Test** with Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 37 Quick Noise Test – Results




Figure 38 Quick Noise Test – Signal

Test Failed

Quick Noise Test Evaluation

Probable cause

- 1** Insufficient lamp warm-up time.
- 2** Absorbing solvent or air bubble in flow cell.
- 3** Dirty or contaminated flow cell.
- 4** Old UV-lamp.

Suggested actions

- Allow detector and UV-lamp turned on for at least 2 hours.
- Ensure the flow cell is filled with water, and free from air bubbles.
- Flush flow cell.
 - Clean the flow cell as described in “[Clean the Max-Light Cartridge Cell](#)” on page 163.
- Exchange the UV-lamp.

ASTM Drift and Noise Test

The ASTM noise test determines the detector noise over a period of 20 minutes. The test is done with installed Max-Light Cartridge Cell or Max-Light Cartridge Test Cell.

This test does also check for the drift. It is also part of the “Self Test” (without checking for the drift).

If the test is performed with the Max-Light Cartridge Test Cell, the test results are not influenced by solvent or pump effects.

When In case of noise and drift problem.

Parts required	#	Description
	1	Max-Light Cartridge Cell (filled with water)
OR	1	Max-Light Cartridge Test Cell

Preparations

- Detector and UV-lamp must be on for at least 2 hours.
- ASTM measurements based on specifications may require longer stabilization times.
- When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.

8 Test Functions and Calibration
ASTM Drift and Noise Test

- 1 Run the **ASTM Drift and Noise Test** with Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 39 ASTM Drift and Noise Test – Results




Figure 40 ASTM Drift and Noise Test – Signal

Test Failed

ASTM Noise Test Evaluation

Probable cause

- 1 Insufficient lamp warm-up time.
- 2 Absorbing solvent or air bubble in flow cell.
- 3 Dirty or contaminated flow cell.
- 4 Old UV-lamp.
- 5 Environment not according to specifications.

Suggested actions

- Allow detector and UV-lamp turned on for at least 2 hours.
- Ensure the flow cell is filled with water, and free from air bubbles.
- Flush flow cell.
 - Clean the flow cell as described in “[Clean the Max-Light Cartridge Cell](#)” on page 163.
- Exchange the UV-lamp.
- Improve environment.

Slit Test

Slit Test (G7117B)

The slit test verifies correct operation of the micromechanical slit.

During the test, the slit is moved through all slit positions while the detector monitors the lamp intensity change. When the slit position is changed, the intensity drop (move to smaller slit) or intensity increase (move to larger slit) must be within a defined range.

If the intensity changes are outside the expected range, the test fails.

When	In case of problems.	
Parts required	#	Description
OR	1	Max-Light Cartridge Cell (filled with water)
	1	Max-Light Cartridge Test Cell
Preparations	• Lamp must be on for at least 10 minutes. • When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.	

- 1 Run the **Slit Test** with the Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 41 Slit Test – Results




Figure 42 Slit Test – Signal

Test Failed

Slit Test Evaluation

Probable cause

- 1** Air bubble in Max-Light Cartridge Cell.
- 2** Old lamp.
- 3** Defective slit assembly.
- 4** Defective detector main board.
- 5** Defective optical unit.

Suggested actions

- Flush the flow cell or use the Max-Light Cartridge Test Cell.
- Run the “Intensity Test”. Exchange the lamp if old or defective.
- Please contact your Agilent service representative.
- Please contact your Agilent service representative.
- Please contact your Agilent service representative.

Slit Test (G7117A)

There is no dedicated slit test for the G7117A DAD FS. To verify the proper function perform the following tests:

- Intensity Test (tests the normal position)
- Dark Current Test (tests the dark position)

Wavelength Verification Test

The detector uses the alpha (656.1 nm) and beta (486 nm) emission lines of the UV-lamp for wavelength calibration. The sharp emission lines enable accurate calibration. When verification is started, the 1-nm slit is moved into the light path automatically. The test is run with the Max-Light Cartridge Cell or with Max-Light Cartridge Test Cell installed.

If the test is performed with the Max-Light Cartridge Test Cell, the test results are not influenced by solvent or pump effects.

When

The detector is calibrated at the factory, and under normal operating conditions should not require recalibration. However, it is advisable to recalibrate:

- after repair of components in the optical unit,
- after exchange of the optical unit or main board,
- after replacing the Max-Light Cartridge Cell or UV-lamp,
- after significant environmental condition changes (temperature, humidity),
- at a regular interval, at least once per year (for example, prior to an Operational Qualification/Performance Verification procedure), and
- when chromatographic results indicate the detector may require recalibration.

Parts required

#	Description
1	Max-Light Cartridge Test Cell or
1	Max-Light Cartridge Cell

Preparations

- Lamp must be on for at least 10 minutes.
- When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.

8 Test Functions and Calibration

Wavelength Verification Test

- 1 Run the **Wavelength Verification Test** with the Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 43 Wavelength Verification – Results

Wavelength Calibration

The detector uses the alpha (656.1 nm) and beta (486 nm) emission lines of the deuterium lamp for wavelength calibration. The sharp emission lines enable more accurate calibration than is possible with holmium oxide. When recalibration is started, the 1 nm slit is moved into the light path automatically (G7117B). The gain is set to zero.

On completion of the scan, the alpha- and beta-line deviations (in nm) are displayed. These values indicate how far the detector calibration deviates from the actual positions of the alpha and beta emission lines. After calibration, the deviation is zero.

To eliminate effects due to absorbing solvents, install the Max-Light Cartridge Test Cell before starting the test.

When	The detector is calibrated at the factory, and under normal operating conditions should not require recalibration. However, it is advisable to recalibrate: • after maintenance (flow cell or UV-lamp), • after repair of components in the optical unit, • after exchange of the optical unit or main board, • after significant environmental condition changes (temperature, humidity), • at a regular interval, at least once per year (for example, prior to an Operational Qualification/Performance Verification procedure), and • when chromatographic results indicate the detector may require recalibration.	
-------------	---	--

Parts required	#	Description
	1	Max-Light Cartridge Test Cell or
	1	Max-Light Cartridge Cell

Preparations	• Detector/lamp must be on for more than 1 hour. • When using a Max-Light Cartridge Cell a flow rate of 0.5 mL/min with water is required.
---------------------	---

NOTE

If the detector is operated in a lab environment that differs at average from the final test environment (25 °C) then the detector should be recalibrated for this temperature.

NOTE

If the detector was repaired (opened covers), the wavelength calibration can be done 10 minutes after lamp on. A final wavelength calibration should be repeated after complete warm-up of the detector.

- 1 Run the **Wavelength Calibration** with the Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 44 Wavelength Calibration - Results

If you select **No**, the test is aborted.

If you select **Yes**, the re-calibration is performed (the offset is corrected).

Wavelength Recalibration Fails

Probable cause	Suggested actions
1 Absorbing solvent or air bubble in Max-Light Cartridge Cell.	Repeat calibration with Max-Light Cartridge Test Cell and compare results.
2 Dirty or contaminated Max-Light Cartridge Cell.	• Ensure the Max-Light Cartridge Cell is filled with water. • Recalibrate.
3 Old UV-lamp.	Exchange the UV-lamp.
4 Dirty or contaminated optical components.	Run the Cell Test. If the test fails, flush the flow cell. See also “Clean the Max-Light Cartridge Cell” on page 163.

NOTE

If the test fails with Max-Light Cartridge Test Cell and new UV-lamp, the optical unit must be replaced.

D/A Converter (DAC) Test

The detector provides analog output of chromatographic signals for use with integrators, chart recorders or data systems. The analog signal is converted from the digital format by the digital-analog-converter (DAC).

The DAC test is used to verify correct operation of the digital-analog-converter by applying a digital test signal to the DAC.

The DAC outputs an analog signal of approximately 50 mV (if the zero offset of the analog output is set to the default value of 5 %) which can be plotted on an integrator. A continuous square wave with an amplitude of 10 μ V and a frequency of approximately 1 cycle/24 seconds is applied to the signal.

The amplitude of the square wave and the peak-to-peak noise are used to evaluate the DAC test.

When

If the analog detector signal is noisy or missing.

Preparations

Lamp must be on for at least 10 minutes. Connect integrator, chart recorder or data system to the detector analog output.

- 1 Run the **D/A Converter (DAC) Test** with the Agilent Lab Advisor (for further information see Online-Help of user interface).




Figure 45 D/A Converter (DAC) Test – Results




Figure 46 D/A Converter (DAC) Test – Example of Integrator Plot

Test Evaluation

The noise on the step should be less than 3 μ V.

Probable cause	Suggested actions
1 Bad cable or grounding problem between detector and external device.	Check or replace the cable.
2 Defective detector main board.	Please contact your Agilent service representative.

Dark Current Test

The dark-current test measures the leakage current from each diode. The test is used to check for leaking diodes which may cause non-linearity at specific wavelengths. During the test, the slit assembly moves to the dark position, cutting off all light falling onto the diode array. Next, the leakage current from each diode is measured, and displayed graphically. The leakage current (represented in counts) for each diode should fall within the limits.

When

In case of problem.

- 1 Run the **Dark Current Test** with the recommended user interface (for further information see Online-Help of user interface).




Figure 47 Dark Current Test – Results

8 Test Functions and Calibration

Dark Current Test




Figure 48 Dark Current Test – Signals

Test Failed

Probable cause

- 1 Defective slit assembly (stray light).
- 2 Defective detector main board.
- 3 Defective PDA/optical unit.

Suggested actions

- Run the "Self-Test" on page 122.
- Run the "Slit Test (G7117B)" on page 136 (part of the "Self-Test" on page 122).

Please contact your Agilent service representative.

Please contact your Agilent service representative.

Other Lab Advisor Functions

EMFs - Early Maintenance Feature

The EMFs screen allows you to view and manage the EMF counters for all modules in all systems.

Filter

☒ All Counters
 ☐ Counters with Limit

		Title	Value	Unit	Limit	Progress	
G7117B G7117B 1290 DAD Serial # PPBAW00058		Accumulated UV Lamp On-Time	3.17	h	0	0%	★
		Number of UV Lamp Ignitions	2	Count	0	0%	★


Activate EMF

Deactivate EMF

Refresh Counters

8 Test Functions and Calibration

Other Lab Advisor Functions



9 Maintenance

Introduction to Maintenance	152
Warnings and Cautions	152
Overview of Maintenance	154
Cleaning the Module	155
Remove and Install Doors	156
Replace the Deuterium Lamp	157
Replace the Max-Light Cartridge Cell	160
Clean the Max-Light Cartridge Cell	163
Storage of Max-Light Cartridge Cell	165
Correcting Leaks	166
Replace Leak Handling System Parts	167
Replace the Module Firmware	169
Information from Module's Assemblies	171

This chapter describes the maintenance of the module.



Introduction to Maintenance

The module is designed for easy maintenance. Maintenance can be done from the front with module in place in the system stack.

NOTE

There are no serviceable parts inside.
Do not open the module.

Warnings and Cautions

WARNING

Toxic, flammable and hazardous solvents, samples and reagents

The handling of solvents, samples and reagents can hold health and safety risks.

- When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.
 - The volume of substances should be reduced to the minimum required for the analysis.
 - Do not operate the instrument in an explosive atmosphere.
-

WARNING

Eye damage by detector light

Eye damage may result from directly viewing the UV-light produced by the lamp of the optical system used in this product.

- Always turn the lamp of the optical system off before removing it.
-

WARNING

Electrical shock

Repair work at the module can lead to personal injuries, e.g. shock hazard, when the cover is opened.

- Do not remove the cover of the module.
 - Only certified persons are authorized to carry out repairs inside the module.
-

WARNING

Personal injury or damage to the product

Agilent is not responsible for any damages caused, in whole or in part, by improper use of the products, unauthorized alterations, adjustments or modifications to the products, failure to comply with procedures in Agilent product user guides, or use of the products in violation of applicable laws, rules or regulations.

- Use your Agilent products only in the manner described in the Agilent product user guides.
-

CAUTION

Safety standards for external equipment

- If you connect external equipment to the instrument, make sure that you only use accessory units tested and approved according to the safety standards appropriate for the type of external equipment.
-

Overview of Maintenance

The following pages describe maintenance (simple repairs) of the detector that can be carried out without opening the main cover.

Table 16 Overview of Maintenance

Procedure	Typical Frequency	Notes
Cleaning of module	If required	
Deuterium lamp exchange	If noise and/or drift exceeds your application limits or lamp does not ignite.	A wavelength calibration test and an intensity test should be performed after replacement.
Flow cell exchange	If leaking or if intensity drops due to contaminated flow cell.	A wavelength calibration test should be performed after replacement.
Leak sensor drying	If leak has occurred.	Check for leaks.
Leak handling System replacement	If broken or corroded.	Check for leaks.

Cleaning the Module

To keep the module case clean, use a soft cloth slightly dampened with water, or a solution of water and mild detergent.

WARNING

Liquid dripping into the electronic compartment of your module can cause shock hazard and damage the module

- Do not use an excessively damp cloth during cleaning.
 - Drain all solvent lines before opening any connections in the flow path.
-


Remove and Install Doors

Parts required	p/n	Description
	5067-5737	Door left
	5067-5736	Door right


NOTE

The figures shown in this procedure exemplarily show the Infinity II Multisampler module.
The principle of how to remove and/or install doors works in the same way for all Infinity II modules.

- 1 Press the release buttons and pull the front door out.



- 2 For the Installation of the front door. Insert the hinges into their guides and move the door in until the release buttons click into their final position.



Replace the Deuterium Lamp

When If noise or drift exceeds application limits or lamp does not ignite.

Tools required **Description**
Screwdriver POZI 1 PT3

Parts required	#	p/n	Description
	1	5190-0917	Long-life Deuterium lamp (8-pin) with RFID tag

Preparations Turn the lamp off.


WARNING

Injury by touching hot lamp
If the detector has been in use, the lamp may be hot.
→ If so, wait for lamp to cool down.


NOTE

The lamp house cover includes a magnet.

1 Open the doors.




2 Locate the lamp cover.




9 Maintenance

Replace the Deuterium Lamp


- 3** Grab the lamp cover and pull it off (it is fixed by a magnet in the bottom of the cover).



- 4** Unplug the lamp connector (1.) and unscrew the two lamp screws (2.) (Pozidriv).




- 5** Remove the lamp and place it on a clean place.




NOTE

Do not touch the glass bulb with your fingers. It may reduce the light output.


- 6** Insert the lamp (RFID tag on the right side).




- 7** Fix the lamp screws (1.) and reconnect the lamp connector (2.).




- 8** Place the lamp cable in the lamp cover.



- 9** Slide the lamp cover into the top position of the metal front (1.) and press the lamp cover completely in until it clicks (2.).



- 10** Close the doors.



- 11** Perform a Wavelength Re-calibration after lamp warm-up.

Replace the Max-Light Cartridge Cell

When If leaking or if intensity drops due to contaminated flow cell.

Tools required **Description**
Wrench, 1/4 inch
for capillary connections

Parts required	p/n	Description
	G4212-60008	Max-Light Cartridge Cell (10 mm, V(σ) 1.0 μL)
	G4212-60007	Max-Light Cartridge Cell (60 mm, V(σ) 4.0 μL)
	G4212-60011	Max-Light Cartridge Test Cell
	G4212-60032	HDR Max-Light Cartridge Cell (3.7 mm, V(σ) 0.4 μL)
	G4212-60038	ULD Max-Light Cartridge Cell (10 mm, V(σ) 0.6 μL)


Preparations Turn the pump off.

NOTE


The flow cell is shipped with a filling of isopropanol. This is to avoid breakage due to subambient conditions. In case the flow cell is not used for some time (stored), then flush the flow cell with iso-propanol.

Remove the Max-Light Cartridge Cell


- 1** Open the doors.




- 2** Disconnect the capillaries from the flow cell cartridge.




- 3** Flip the cartridge lever towards the front (down).



- 4** Pull the cartridge holder completely out.








- 5** Remove the flow cell cartridge.



9 Maintenance

Replace the Max-Light Cartridge Cell

Install the Max-Light Cartridge Cell

1 Insert the flow cell cartridge. 	2 Slide the cartridge holder completely into the detector. 
3 Lift the cartridge lever up into the final position. 	4 Reconnect the capillaries. 
5 Close the doors. 	

Clean the Max-Light Cartridge Cell

When Low counts on Intensity Test or Cell Test (failed tests)

Tools required	p/n	Description
		Alcohol (Iso-propanol or Ethanol)
		Lens tissue or Q-tips [®]
	5190-0530	Cell cleaning solvent

- 1 Flush the flow cell with the alcohol for some time.
- 2 Remove the cell from the cartridge holder (see [“Replace the Max-Light Cartridge Cell”](#) on page 160).
- 3 Carefully clean the light inlet and outlet of the cell using lens tissue or Q-tips[®] with alcohol.

NOTE

If Q-tips[®] are used, ensure that no cotton fluff remains at the inlet or outlet.

NOTE

Do not touch the light inlet and outlet of the cell with your fingers. This will add a layer of contamination on the window and reduce the light throughput.

- 4 Flush the flow cell with water and repeat the Intensity Test and or Cell Test.
- 5 If the cleaning with the alcohol did not improve the results, the flow cell might be cleaned with cleaning fluid (PN 5190-0530). Use a concentration of 0.5 – 2 v/v % (cleaning fluid/water). Use a syringe to fill the flow cell with cleaning fluid.

The following cleaning protocols are recommended:

- Maximum 3 hours at 25 – 30 °C or
- 30 – 40 min at 30 – 35 °C.

NOTE

The optimal concentration depends on the water quality, the contamination, the temperature, and other factors. The use of demineralized water may improve the cleaning characteristics.

- 6** Repeat step 4 on page 163.
- 7** If tests fail again, the flow cell might be replaced if the chromatographic performance cannot be accepted.

Storage of Max-Light Cartridge Cell




- 1 Flush the Max- Light Cartridge Flow Cell with iso- propanol or methanol and insert the plugs into the cell inlet and outlet (see [“Replace the Max-Light Cartridge Cell”](#) on page 160).
- 2 Remove the Max-Light Cartridge Cell from the cartridge holder of the detector.
- 3 Replace the black hoods, that secure the cell light inlet and outlet.
- 4 Store the Max-Light Cartridge Cell in plastic case provided with the Max-Light Cartridge Flow Cell.

Correcting Leaks

When If leak has occurred.

Tools required **Description**
Tissue


Preparations Turn the pump off.

1 Open the doors. 	2 Use tissue to dry the leak sensor area (1). Observe the capillary connections and the flow cell area (2) for leaks and correct, if required. 
3 Close the doors. 	


Replace Leak Handling System Parts

Parts required	p/n	Description
	5043-0856	Leak Adapter
	5063-6527	Tubing assembly, i.d. 6 mm, o.d. 9 mm, 1.2 m (to waste) approximately 85 mm required


1 Open the doors.




2 Locate the Leak Adapter (1) and Tubing (2).



3 Press the Leak Adapter down and remove it together with the tubing.




4 Install the Leak Adapter by pressing it into the Main Cover.




9 Maintenance

Replace Leak Handling System Parts

- 5** Insert the Tubing (approximately 85 mm required for replacement) between Leak Adapter outlet and Leak Panel.



- 6** Close the doors.



Replace the Module Firmware

When	The installation of newer firmware might be necessary • if a newer version solves problems of older versions or • to keep all systems on the same (validated) revision. The installation of older firmware might be necessary • to keep all systems on the same (validated) revision or • if a new module with newer firmware is added to a system or • if third party control software requires a special version.
-------------	--

Tools required	Description
	Agilent Lab Advisor software
OR	Instant Pilot G4208A (only if supported by module)

Parts required	#	Description
	1	Firmware, tools and documentation from Agilent web site

Preparations	Read update documentation provided with the Firmware Update Tool.
---------------------	---

To upgrade/downgrade the module's firmware carry out the following steps:

- 1 Download the required module firmware, the latest LAN/USB FW Update Tool and the documentation from the Agilent web.
http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?wid=69761
- 2 For loading the firmware into the module follow the instructions in the documentation.

Module Specific Information

Table 17 Module Specific Information (G7117A/G7117B)

G7117B 1290 DAD		G7117A 1290 DAD FS
Initial firmware (main and resident)		B.06.70
Compatibility with 1100/1200/1260/1290 series modules	When using the detector in a system, all other modules must have firmware from set 6.50 (latest version) or later (main and resident). Otherwise the communication will not work.	
Conversion to / emulation	G4212A	G4212B

Information from Module's Assemblies

Lamp and Flow Cell RFID Tag

The detector is equipped with a UV lamp and flow cell identification system using RFID (radio frequency identification) tags attached to the assemblies and RFID tag readers at the optical unit. The table below lists all parameters stored in the RFID tag.

Table 18 RFID Tag Data

Lamp information	Flow cell information
• product number	• product number
• serial number	• serial number
• production date	• production date
• accumulated UV on time (in hours)	• nominal path length of the cell (in mm)
• actual UV lamp on time (in hours)	• cell volume (σ) in μL
• number of ignitions	• maximum pressure (in bar)
• date of last intensity test	• date of last cell test

NOTE


The pressure value is always displayed in bar, even if the user interface uses other units, e.g. PSI.

Serial Number and Firmware Revision

The user interface provides module specific information that is stored in the main board. These are for example the serial number, firmware revision.

9 Maintenance

Information from Module's Assemblies



10

Parts and Materials for Maintenance

Overview of Maintenance Parts [174](#)

Accessory Kit [176](#)

This chapter provides information on parts for maintenance.



Overview of Maintenance Parts





Figure 49 Overview of maintenance parts

Item	p/n	Description
	1	5190-0917 Long-life Deuterium lamp (8-pin) with RFID tag
	2	G4212-60008 Max-Light Cartridge Cell (10 mm, V(σ) 1.0 μ L)
OR	2	G4212-60007 Max-Light Cartridge Cell (60 mm, V(σ) 4.0 μ L)
OR	2	G4212-60011 Max-Light Cartridge Test Cell
OR	2	G4212-60032 HDR Max-Light Cartridge Cell (3.7 mm, V(σ) 0.4 μ L)
OR	2	G4212-60038 ULD Max-Light Cartridge Cell (10 mm, V(σ) 0.6 μ L)
	3	5043-0856 Leak Adapter
	4	5063-6527 Tubing assembly, i.d. 6 mm, o.d. 9 mm, 1.2 m (to waste) for Waste and Leak Adapter (ca. 85 mm required)
		5062-8535 Waste accessory kit (Flow Cell to waste)
	5	5067-5737 Door left
	6	5067-5736 Door right
	7	5043-1013 Tubing Clip

For cables, see [“Cable Overview”](#) on page 178.

Accessory Kit

G7117-68755		
Accessory Kit	p/n	Description
	5062-8535	Waste accessory kit
	5063-6527	Tubing assembly, i.d. 6 mm, o.d. 9 mm, 1.2 m (to waste) see item 4 in Figure 49 on page 174
	5181-1516	CAN cable, Agilent module to module, 0.5 m
	5500-1155	Tube Connector, 90 degree, ID 6.4
	5043-1013	Tubing Clip see item 7 in Figure 49 on page 174



11 Identifying Cables

Cable Overview	178
Analog Cables	180
Remote Cables	182
CAN/LAN Cables	186
RS-232 Cables	187
USB Cables	187

This chapter provides information on cables used with the Agilent 1200 Infinity Series modules.



Cable Overview

NOTE

Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

Analog cables

p/n	Description
35900-60750	Agilent 35900A A/D converter
01046-60105	Analog cable (BNC to general purpose, spade lugs)

Remote cables

p/n	Description
5188-8029	ERI to general purpose
5188-8044	Remote Cable ERI – ERI
5188-8045	Remote Cable APG – ERI
5061-3378	Remote Cable to 35900 A/D converter
01046-60201	Agilent module to general purpose

CAN cables

p/n	Description
5181-1516	CAN cable, Agilent module to module, 0.5 m
5181-1519	CAN cable, Agilent module to module, 1 m

LAN cables

p/n	Description
5023-0203	Cross-over network cable, shielded, 3 m (for point to point connection)
5023-0202	Twisted pair network cable, shielded, 7 m (for point to point connection)

RS-232 cables (not for FUSION board)

p/n	Description
G1530-60600	RS-232 cable, 2 m
RS232-61601	RS-232 cable, 2.5 m Instrument to PC, 9-to-9 pin (female). This cable has special pin-out, and is not compatible with connecting printers and plotters. It's also called "Null Modem Cable" with full handshaking where the wiring is made between pins 1-1, 2-3, 3-2, 4-6, 5-5, 6-4, 7-8, 8-7, 9-9.
5181-1561	RS-232 cable, 8 m

USB cables


p/n	Description
5188-8050	USB A M-USB Mini B 3 m (PC-Module)
5188-8049	USB A F-USB Mini B M OTG (Module to Flash Drive)

Analog Cables



One end of these cables provides a BNC connector to be connected to Agilent modules. The other end depends on the instrument to which connection is being made.


Agilent Module to 35900 A/D converters

p/n 35900-60750	35900	Pin Agilent module	Signal Name
	1		Not connected
	2	Shield	Analog -
	3	Center	Analog +

Agilent Module to BNC Connector

p/n 8120-1840	Pin BNC	Pin Agilent module	Signal Name
	Shield	Shield	Analog -
	Center	Center	Analog +

Agilent Module to General Purpose

p/n 01046-60105	Pin	Pin Agilent module	Signal Name
	1		Not connected
	2	Black	Analog -
	3	Red	Analog +

Remote Cables

ERI (Enhanced Remote Interface)

5188-8029 ERI to general purpose

p/n 5188-8029	pin	Enhanced Remote	Classic Remote
D-Sub female 15way user's view to connector	1	I01	START REQUEST
	2	I02	STOP
	3	I03	READY
	4	I04	POWER ON
	5	I05	NOT USED
	6	I06	SHUT DOWN
	7	I07	START
	8	I08	PREPARE
	9	1wire DATA	
	10	DGND	
	11	+5V ERI out	
	12	PGND	
	13	PGND	
	14	+24V ERI out	
	15	+24V ERI out	

5188-8044 ERI to ERI (Connector D_Subminiature 15 pin)

Table 19 5188-8044 ERI to ERI

p/n 5188-8044	Pin (ERI)	Signal	Pin (ERI)
	10	GND	10
	1	Start Request	1
	2	Stop	2
	3	Ready	3
	5	Power on	5
	4	Future	4
	6	Shut Down	6
	7	Start	7
	8	Prepare	8
	Ground Connection	Cable Shielding	NC

5188-8045 ERI to APG (Connector D_Subminiature 15 pin (ERI), Connector D_Subminiature 9 pin (APG))


p/n 5188-8045	Pin (ERI)	Signal	Pin (APG)
	10	GND	1
	1	Start Request	9
	2	Stop	8
	3	Ready	7
	5	Power on	6
	4	Future	5
	6	Shut Down	4
	7	Start	3
	8	Prepare	2
	Ground Connection	Cable Shielding	NC

11 Identifying Cables
Remote Cables




One end of these cables provides a Agilent Technologies APG (Analytical Products Group) remote connector to be connected to Agilent modules. The other end depends on the instrument to be connected to.


Agilent Module to Agilent 35900 A/D Converters

p/n 5061-3378	Pin 35900 A/D	Pin Agilent module	Signal Name	Active (TTL)
	1 - White	1 - White	Digital ground	
	2 - Brown	2 - Brown	Prepare run	Low
	3 - Gray	3 - Gray	Start	Low
	4 - Blue	4 - Blue	Shut down	Low
	5 - Pink	5 - Pink	Not connected	
	6 - Yellow	6 - Yellow	Power on	High
	7 - Red	7 - Red	Ready	High
	8 - Green	8 - Green	Stop	Low
	9 - Black	9 - Black	Start request	Low

Agilent Module to General Purpose

p/n 01046-60201	Wire Color	Pin Agilent module	Signal Name	Active (TTL)
	White	1	Digital ground	
	Brown	2	Prepare run	Low
	Gray	3	Start	Low
	Blue	4	Shut down	Low
	Pink	5	Not connected	
	Yellow	6	Power on	High
	Red	7	Ready	High
	Green	8	Stop	Low
	Black	9	Start request	Low

CAN/LAN Cables



Both ends of this cable provide a modular plug to be connected to Agilent modules CAN or LAN connectors.

CAN Cables

p/n	Description
5181-1516	CAN cable, Agilent module to module, 0.5 m
5181-1519	CAN cable, Agilent module to module, 1 m

LAN Cables

p/n	Description
5023-0203	Cross-over network cable, shielded, 3 m (for point to point connection)
5023-0202	Twisted pair network cable, shielded, 7 m (for point to point connection)

RS-232 Cables

p/n	Description
G1530-60600	RS-232 cable, 2 m
RS232-61601	RS-232 cable, 2.5 m Instrument to PC, 9-to-9 pin (female). This cable has special pin-out, and is not compatible with connecting printers and plotters. It's also called "Null Modem Cable" with full handshaking where the wiring is made between pins 1-1, 2-3, 3-2, 4-6, 5-5, 6-4, 7-8, 8-7, 9-9.
5181-1561	RS-232 cable, 8 m


USB Cables

To connect a USB Flash Drive use a USB OTG cable with Mini-B plug and A socket.

p/n	Description
5188-8050	USB A M-USB Mini B 3 m (PC-Module)
5188-8049	USB A F-USB Mini B M OTG (Module to Flash Drive)

11 Identifying Cables

USB Cables



12 Hardware Information

Firmware Description	190
Electrical Connections	193
Rear view of the module	194
Information on Instrument Serial Number	194
Interfaces	196
Overview Interfaces	197
ERI (Enhanced Remote Interface)	199
USB (Universal Serial Bus)	202
Setting the 6-bit Configuration Switch	203
Special Settings	205
Instrument Layout	207
Early Maintenance Feedback	208

This chapter describes the detector in more detail on hardware and electronics.



Firmware Description

The firmware of the instrument consists of two independent sections:

- a non-instrument specific section, called *resident system*
- an instrument specific section, called *main system*

Resident System

This resident section of the firmware is identical for all Agilent 1100/1200/1220/1260/1290 series modules. Its properties are:

- the complete communication capabilities (CAN, LAN, USB and RS-232C)
- memory management
- ability to update the firmware of the 'main system'

Main System

Its properties are:

- the complete communication capabilities (CAN, LAN, USB and RS-232C)
- memory management
- ability to update the firmware of the 'resident system'

In addition the main system comprises the instrument functions that are divided into common functions like

- run synchronization through APG remote,
- error handling,
- diagnostic functions,
- or module specific functions like
 - internal events such as lamp control, filter movements,
 - raw data collection and conversion to absorbance.

Firmware Updates

Firmware updates can be done using the following tools (latest version should be used):

- Agilent Lab Advisor software with files on the hard disk (*)
- Firmware Update Tool with local files on the hard disk (*)
- Instant Pilot (G4208A) with files on a USB Flash Disk

(*) Required tools, firmware and documentation are available from the Agilent web:

http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?whid=69761

The file naming conventions are:

PPPP_RVVV_XXX.dlb, where

PPPP is the product number, for example, 1315B for the G1315B DAD,

R the firmware revision, for example, A for G1315B or B for the G1315C DAD,

VVV is the revision number, for example 650 is revision 6.50,

XXX is the build number of the firmware.

For instructions on firmware updates refer to section *Replacing Firmware* in chapter "Maintenance" or use the documentation provided with the *Firmware Update Tools*.

NOTE

Update of main system can be done in the resident system only. Update of the resident system can be done in the main system only.

Main and resident firmware must be from the same set.




Figure 50 Firmware Update Mechanism

NOTE

Some modules are limited in downgrading due to their main board version or their initial firmware revision. For example, a G1315C DAD SL cannot be downgraded below firmware revision B.01.02 or to a A.xx.xx.

Some modules can be re-branded (e.g. G1314C to G1314B) to allow operation in specific control software environments. In this case the feature set of the target type are use and the feature set of the original are lost. After re-branding (e.g. from G1314B to G1314C), the original feature set is available again.

All these specific informations are described in the documentation provided with the firmware update tools.

The firmware update tools, firmware and documentation are available from the Agilent web.

- http://www.chem.agilent.com/_layouts/agilent/downloadFirmware.aspx?whid=69761

Electrical Connections

- The CAN bus is a serial bus with high-speed data transfer. The two connectors for the CAN bus are used for internal module data transfer and synchronization.
- One analog output provides signals for integrators or data handling systems.
- The ERI/REMOTE connector may be used in combination with other analytical instruments from Agilent Technologies if you want to use features such as start, stop, common shutdown, prepare, and so on.
- With the appropriate software, the LAN connector may be used to control the module from a computer through a LAN connection. This connector is activated and can be configured with the configuration switch.
- With the appropriate software, the USB connector may be used to control the module from a computer through a USB connection.
- The power input socket accepts a line voltage of 100 – 240 VAC $\pm$ 10 % with a line frequency of 50 or 60 Hz. Maximum power consumption varies by module. There is no voltage selector on your module because the power supply has wide-ranging capability. There are no externally accessible fuses because automatic electronic fuses are implemented in the power supply.

NOTE

Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.

Rear view of the module




Figure 51 Rear view of detector (example shows a G7114A/B VWD) – electrical connections and label

Information on Instrument Serial Number

Serial Number Information 1200 Series and 1290 Infinity

The serial number information on the instrument labels provide the following information:

CCYWWSSSSS	Format
CC	country of manufacturing • DE = Germany • JP = Japan • CN = China
YWW	year and week of last major manufacturing change, e.g. 820 could be week 20 of 1998 or 2008
SSSSS	real serial number

Serial Number Information 1260 Infinity

The serial number information on the instrument labels provide the following information:

CCXZZ00000	Format
CC	Country of manufacturing • DE = Germany • JP = Japan • CN = China
X	Alphabetic character A-Z (used by manufacturing)
ZZ	Alpha-numeric code 0-9, A-Z, where each combination unambiguously denotes a module (there can be more than one code for the same module)
00000	Serial number

Interfaces

The Agilent 1200 Infinity Series II modules provide the following interfaces:

Table 20 Agilent 1200 Infinity II Series Interfaces

Module	CAN	USB	LAN (on-board)	RS-232	Analog	APG (A) / ERI (E)	Special
Pumps							
G7104A Flexible Pump	2	No	Yes	Yes	1	A	
G7120A High Speed Pump	2	No	Yes	Yes	1	A	
Samplers							
G7129A/B ALS	2	Yes	Yes	No	No	E	
G7167A/B Multisampler	2	Yes	Yes	No	No	E	
Detectors							
G7114A/B VWD	2	Yes	Yes	No	1	E	
G7117A/B DAD	2	Yes	Yes	No	1	E	
G7115A/B DAD	2	Yes	Yes	No	1	E	
Others							
G7116B MCT	2	No	No	No	No	No	Requires a HOST module via CAN

NOTE

The detector (DAD/MWD/FLD/VWD/RID) is the preferred access point for control via LAN. The inter-module communication is done via CAN.

- CAN connectors as interface to other modules
- LAN connector as interface to the control software
- RS-232C as interface to a computer
- USB (Universal Series Bus) as interface to a computer
- REMOTE connector as interface to other Agilent products
- Analog output connector(s) for signal output

Overview Interfaces

CAN

The CAN is inter-module communication interface. It is a 2-wire serial bus system supporting high speed data communication and real-time requirement.

LAN

The modules have either an interface slot for an LAN card (e.g. Agilent G1369B/C LAN Interface) or they have an on-board LAN interface (e.g. detectors G1315C/D DAD and G1365C/D MWD). This interface allows the control of the module/system via a PC with the appropriate control software. Some modules have neither on-board LAN nor an interface slot for a LAN card (e.g. G1170A Valve Drive or G4227A Flex Cube). These are hosted modules and require a Host module with firmware B.06.40 or later or with additional G1369C LAN Card.

NOTE

If an Agilent detector (DAD/MWD/FLD/VWD/RID) is in the system, the LAN should be connected to the DAD/MWD/FLD/VWD/RID (due to higher data load). If no Agilent detector is part of the system, the LAN interface should be installed in the pump or autosampler.

USB

The USB interface replaces the RS-232 Serial interface in new FUSION generation modules. For details on USB refer to “[USB \(Universal Serial Bus\)](#)” on page 202.

Analog Signal Output

The analog signal output can be distributed to a recording device. For details refer to the description of the module's main board.

Remote (ERI)

The ERI (Enhanced Remote Interface) connector may be used in combination with other analytical instruments from Agilent Technologies if you want to use features as common shut down, prepare, and so on.

It allows easy connection between single instruments or systems to ensure coordinated analysis with simple coupling requirements.

The subminiature D connector is used. The module provides one remote connector which is inputs/outputs (wired- or technique).

To provide maximum safety within a distributed analysis system, one line is dedicated to **SHUT DOWN** the system's critical parts in case any module detects a serious problem. To detect whether all participating modules are switched on or properly powered, one line is defined to summarize the **POWER ON** state of all connected modules. Control of analysis is maintained by signal readiness **READY** for next analysis, followed by **START** of run and optional **STOP** of run triggered on the respective lines. In addition **PREPARE** and **START REQUEST** may be issued. The signal levels are defined as:

- standard TTL levels (0 V is logic true, + 5.0 V is false),
- fan-out is 10,
- input load is 2.2 kOhm against + 5.0 V, and
- output are open collector type, inputs/outputs (wired- or technique).

NOTE

All common TTL circuits operate with a 5 V power supply. A TTL signal is defined as "low" or L when between 0 V and 0.8 V and "high" or H when between 2.0 V and 5.0 V (with respect to the ground terminal).

Table 21 ERI signal distribution

Pin	Signal	Description
1	START REQUEST	(L) Request to start injection cycle (for example, by start key on any module). Receiver is the autosampler.
2	STOP	(L) Request to reach system ready state as soon as possible (for example, stop run, abort or finish and stop injection). Receiver is any module performing run-time controlled activities.
3	READY	(H) System is ready for next analysis. Receiver is any sequence controller.
4	POWER ON	(H) All modules connected to system are switched on. Receiver is any module relying on operation of others.
5		Not used
6	SHUT DOWN	(L) System has serious problem (for example, leak: stops pump). Receiver is any module capable to reduce safety risk.
7	START	(L) Request to start run / timetable. Receiver is any module performing run-time controlled activities.
8	PREPARE	(L) Request to prepare for analysis (for example, calibration, detector lamp on). Receiver is any module performing pre-analysis activities.

Special Interfaces

There is no special interface for this module.

ERI (Enhanced Remote Interface)

ERI replaces the AGP Remote Interface that is used in the HP 1090/1040/1050/1100 HPLC systems and Agilent 1100/1200/1200 Infinity HPLC modules. All new 1200 Infinity II products using the FUSION core electronics use ERI. This interface is already used in the Agilent Universal Interface Box 2 (UIB2)

ERI Description

The ERI interface contains eight individual programmable input/output pins. In addition, it provides 24 V power and 5 V power and a serial data line to detect and recognize further add-ons that could be connected to this interface. This way the interface can support various additional devices like sensors, triggers (in and out) and small controllers, etc.





Figure 52 Location of the ERI interface (example shows a G7114A/B VWD)



IO (Input/Output) Lines

- Eight generic bi-directional channels (input or output).
- Same as the APG Remote.
- Devices like valves, relays, ADCs, DACs, controllers can be supported/controlled.

1-Wire Data (Future Use)

This serial line can be used to read out an EPROM or write into an EPROM of a connected ERI-device. The firmware can detect the connected type of device automatically and update information in the device (if required).

5V Distribution (Future Use)

- Available directly after turn on of the hosting module (assures that certain base functionality of the device can be detected by firmware).
- For digital circuits or similar.
- Provided 500 mA maximum.
- Short-circuit proof with automatic switch off (by firmware).

24V Distribution (Future Use)

- Available by firmware command (defined turn on/off).
- For devices that need higher power
 - Class 0: 0.5 A maximum (12 W)
 - Class 1: 1.0 A maximum (24 W)
 - Class 2: 2.0 A maximum (48 W)
- Class depends on hosting module's internal power overhead.
- If a connected device requires more power the firmware detects this (overcurrent detection) and provides the information to the user interface.
- Fuse used for safety protection (on board).
- Short circuit will be detected through hardware.

USB (Universal Serial Bus)

USB (Universal Serial Bus) - replaces RS232, supports:

- a PC with control software (for example Agilent Lab Advisor)
- USB Flash Disk

Setting the 6-bit Configuration Switch

The 6-bit configuration switch is located at the rear of the module with FUSION electronics. Switch settings provide configuration parameters for LAN and instrument specific initialization procedures.

All modules with FUSION electronics:

- Default is ALL switches DOWN (best settings).
 - Default IP address for LAN 192.168.254.11
- For specific LAN modes switches 4-5 must be set as required.
- For boot resident/cold start modes switches 1+2 or 6 must be UP.




Figure 53 Location of Configuration switch (example shows a G7114A/B VWD)

12 Hardware Information

Setting the 6-bit Configuration Switch

Table 22 6-bit Configuration Switch

	Mode	Function/Setting				
	Switch 1	Switch 2	Switch 3	Switch 4	Switch 5	Switch 6
COM ¹	0	n.a. ²	n.a.	LAN Init Mode		n.a.
Use Default IP Address ³		0	0	0	0	0
Use Stored IP Address		0	0	0	1	0
Use DHCP to request IP Address ⁴		0	0	1	0	0
Test	1	System	n.a.	n.a.	n.a.	ColdStart
Boot Main System / Keep Data		0	0	0	0	0
Boot Resident System / Keep Data		1	0	0	0	0
Boot Main System / Revert to Default Data		0	0	0	0	1
Boot Resident System / Revert to Default Data		1	0	0	0	1

¹ When selecting mode COM, settings are stored to non-volatile memory. When selecting mode TEST, COM settings are taken from non-volatile memory.

² not assigned - Always keep these switches on position '0' (off)

³ Default IP Address is 192.168.254.11

⁴ Host Name will be the MAC address.

Special Settings

Boot-Resident/Main

Firmware update procedures may require this mode in case of firmware loading errors (main/resident firmware part).

If you use the following switch settings and power the instrument up again, the instrument firmware stays in the resident/main mode. In resident mode, it is not operable as a module. It only uses basic functions of the operating system for example, for communication. In this mode the main firmware can be loaded (using update utilities).

Forced Cold Start

A forced cold start can be used to bring the module into a defined mode with default parameter settings.

- Boot Main System / Revert to Default Data

The instrument will boot to main mode and changes to the module's default parameter. May be also required to load resident firmware into the module.

- Boot Resident System / Revert to Default Data

The instrument will boot to resident mode and changes to the module's default parameter. May be also required to load main firmware into the module.

CAUTION

Loss of data

Forced cold start erases all methods and data stored in the non-volatile memory. Exceptions are calibration settings, diagnosis and repair log books which will not be erased.

→ Save your methods and data before executing a forced cold start.

If you use the following switch settings and power the instrument up again, it will start as described above.

12 Hardware Information

Setting the 6-bit Configuration Switch

Table 23 Boot Resident / Forced Coldstart

	SW1	SW2	SW3	SW4	SW5	SW6	Init Mode
	1	0	0	0	0	0	Boot Main System / Keep Data
	1	1	0	0	0	0	Boot Resident System / Keep Data
	1	0	0	0	0	1	Boot Main System / Revert to Default Data
	1	1	0	0	0	1	Boot Resident System / Revert to Default Data
Note: The setting '0' (down) is essential.							

Instrument Layout

The industrial design of the module incorporates several innovative features. It uses Agilent's E-PAC concept for the packaging of electronics and mechanical assemblies. This concept is based upon the use of expanded polypropylene (EPP) layers of foam plastic spacers in which the mechanical and electronic boards components of the module are placed. This pack is then housed in a metal inner cabinet which is enclosed by a plastic external cabinet. The advantages of this packaging technology are:

- virtual elimination of fixing screws, bolts or ties, reducing the number of components and increasing the speed of assembly/disassembly,
- the plastic layers have air channels molded into them so that cooling air can be guided exactly to the required locations,
- the plastic layers help cushion the electronic and mechanical parts from physical shock, and
- the metal inner cabinet shields the internal electronics from electromagnetic interference and also helps to reduce or eliminate radio frequency emissions from the instrument itself.

Early Maintenance Feedback

Maintenance requires the exchange of components which are subject to wear or stress. Ideally, the frequency at which components are exchanged should be based on the intensity of usage of the module and the analytical conditions, and not on a predefined time interval. The early maintenance feedback (**EMF**) feature monitors the usage of specific components in the instrument, and provides feedback when the user-selectable limits have been exceeded. The visual feedback in the user interface provides an indication that maintenance procedures should be scheduled.

EMF Counters

EMF counters increment with use and can be assigned a maximum limit which provides visual feedback in the user interface when the limit is exceeded. Some counters can be reset to zero after the required maintenance procedure.

Lamp Type	Counter Reset	Comment
lamp with RFID tag	NO	
lamp without RFID tag	YES	via Lab Advisor or Instant Pilot

The detector provides the following EMF counters:

- Deuterium Lamp On-Time
- Number of UV lamp ignitions

Using the EMF Counters

The user-settable **EMF** limits for the **EMF Counters** enable the early maintenance feedback to be adapted to specific user requirements. The useful maintenance cycle is dependent on the requirements for use. Therefore, the definition of the maximum limits need to be determined based on the specific operating conditions of the instrument.


Setting the EMF Limits

The setting of the **EMF** limits must be optimized over one or two maintenance cycles. Initially the default **EMF** limits should be set. When instrument performance indicates maintenance is necessary, take note of the values displayed by the **EMF counters**. Enter these values (or values slightly less than the displayed values) as **EMF** limits, and then reset the **EMF counters** to zero. The next time the **EMF counters** exceed the new **EMF** limits, the **EMF** flag will be displayed, providing a reminder that maintenance needs to be scheduled.

NOTE

This function is only available via Agilent Lab Advisor or Instant Pilot.

12 Hardware Information
Early Maintenance Feedback



13 LAN Configuration

What You Have to Do First	212
TCP/IP parameter configuration	213
Configuration Switches	214
Initialization Mode Selection	215
Dynamic Host Configuration Protocol (DHCP)	219
General Information (DHCP)	219
Setup (DHCP)	220
Manual Configuration	222
With Telnet	223
PC and Agilent ChemStation Setup	226
PC Setup for Local Configuration	226
Agilent ChemStation Setup	229

This chapter provides information on connecting the module to the Agilent ChemStation PC.



What You Have to Do First

The module has an on- board LAN communication interface.

NOTE

This chapter is generic and may show figures that differ from your module. The functionality is the same.

- 1 Note the MAC (Media Access Control) address for further reference. The MAC or hardware address of the LAN interfaces is a world wide unique identifier. No other network device will have the same hardware address. The MAC address can be found on a label at the rear of the module underneath the configuration switch (see [Figure 55](#) on page 212).




Figure 54 MAC-Label

- 2 Connect the instrument's LAN interface (see [Figure 55](#) on page 212) to
 - the PC network card using a crossover network cable (point-to-point) or
 - a hub or switch using a standard LAN cable.




Figure 55 Location of LAN interfaces and MAC label

TCP/IP parameter configuration

To operate properly in a network environment, the LAN interface must be configured with valid TCP/IP network parameters. These parameters are:

- IP address
- Subnet Mask
- Default Gateway

The TCP/IP parameters can be configured by the following methods:

- by automatically requesting the parameters from a network-based BOOTP Server (using the so-called Bootstrap Protocol)
- by automatically requesting the parameters from a network-based DHCP Server (using the so-called Dynamic Host Configuration Protocol). This mode requires a LAN-onboard Module or a G1369C LAN Interface card, see [“Setup \(DHCP\)”](#) on page 220
- by manually setting the parameters using Telnet

The LAN interface differentiates between several initialization modes. The initialization mode (short form ‘init mode’) defines how to determine the active TCP/IP parameters after power-on. The parameters may be derived from a Bootp cycle, non-volatile memory or initialized with known default values. The initialization mode is selected by the configuration switch, see [Table 25](#) on page 215.

Configuration Switches

The configuration switch can be accessed at the rear of the module.




Figure 56 Location of Configuration switch (example shows a G7114A/B VWD)

The module is shipped with all switches set to OFF, as shown above.

NOTE

To perform any LAN configuration, SW1 and SW2 must be set to OFF.

Table 24 Factory Default Settings

Initialization ('Init') Mode	Bootp, all switches down. For details see "Initialization Mode Selection" on page 215
------------------------------	---

Initialization Mode Selection

The following initialization (init) modes are selectable:

Table 25 Initialization Mode Switches

	SW1	SW2	SW3	SW4	SW5	SW6	Init Mode
	0	0	0	0	0	0	Use Default IP Address
	0	0	0	0	1	0	Use Stored IP Address
	0	0	0	1	0	0	Use DHCP

Note: The setting '0' (down) is essential.

Default IP address for LAN is 192.168.254.11.

DHCP address is the module's LAN MAC address.

Bootp

When the initialization mode **Bootp** is selected, the module tries to download the parameters from a **Bootp** Server. The parameters obtained become the active parameters immediately. They are not stored to the non-volatile memory of the module. Therefore, the parameters are lost with the next power cycle of the module.




Figure 57 Bootp (Principle)

Bootp & Store

When **Bootp & Store** is selected, the parameters obtained from a **Bootp** Server become the active parameters immediately. In addition, they are stored to the non-volatile memory of the module. Thus, after a power cycle they are still available. This enables a kind of bootp once configuration of the module.

Example: The user may not want to have a **Bootp** Server be active in his network all the time. But on the other side, he may not have any other configuration method than **Bootp**. In this case he starts the **Bootp** Server temporarily, powers on the module using the initialization mode **Bootp & Store**, waits for the **Bootp** cycle to be completed, closes the **Bootp** Server and powers off the module. Then he selects the initialization mode Using Stored and powers on the module again. From now on, he is able to establish the TCP/IP connection to the module with the parameters obtained in that single **Bootp** cycle.




Figure 58 Bootp & Store (Principle)

NOTE

Use the initialization mode **Bootp & Store** carefully, because writing to the non-volatile memory takes time. Therefore, when the module shall obtain its parameters from a **Bootp** Server every time it is powered on, the recommended initialization mode is **Bootp**!

Using Stored

When initialization mode **Using Stored** is selected, the parameters are taken from the non-volatile memory of the module. The TCP/IP connection will be established using these parameters. The parameters were configured previously by one of the described methods.




Figure 59 Using Stored (Principle)

Using Default

When **Using Default** is selected, the factory default parameters are taken instead. These parameters enable a TCP/IP connection to the LAN interface without further configuration, see [Table 26](#) on page 218.




Figure 60 Using Default (Principle)

NOTE

Using the default address in your local area network may result in network problems. Take care and change it to a valid address immediately.

Table 26 Using Default Parameters

IP address:	192.168.254.11
Subnet Mask:	255.255.255.0
Default Gateway	not specified

Since the default IP address is a so-called local address, it will not be routed by any network device. Thus, the PC and the module must reside in the same subnet.

The user may open a Telnet session using the default IP address and change the parameters stored in the non-volatile memory of the module. He may then close the session, select the initialization mode Using Stored, power-on again and establish the TCP/IP connection using the new parameters.

When the module is wired to the PC directly (e.g. using a cross-over cable or a local hub), separated from the local area network, the user may simply keep the default parameters to establish the TCP/IP connection.

NOTE

In the **Using Default** mode, the parameters stored in the memory of the module are not cleared automatically. If not changed by the user, they are still available, when switching back to the mode Using Stored.

Dynamic Host Configuration Protocol (DHCP)

General Information (DHCP)

The Dynamic Host Configuration Protocol (DHCP) is an auto configuration protocol used on IP networks. The DHCP functionality is available on all Agilent HPLC modules with on-board LAN Interface or LAN Interface Card, and “B”-firmware (B.06.40 or above).

When the initialization mode “DHCP” is selected, the card tries to download the parameters from a DHCP Server. The parameters obtained become the active parameters immediately. They are not stored to the non-volatile memory of the card.

Besides requesting the network parameters, the card also submits its hostname to the DHCP Server. The hostname equals the MAC address of the card, e.g. *0030d3177321*. It is the DHCP server's responsibility to forward the hostname/address information to the Domain Name Server. The card does not offer any services for hostname resolution (e.g. NetBIOS).




Figure 61 DHCP (Principle)

NOTE

- 1 It may take some time until the DHCP server has updated the DNS server with the hostname information.
- 2 It may be necessary to fully qualify the hostname with the DNS suffix, e.g. *0030d3177321.country.company.com*.
- 3 The DHCP server may reject the hostname proposed by the card and assign a name following local naming conventions.

13 LAN Configuration

Dynamic Host Configuration Protocol (DHCP)

Setup (DHCP)

Software required

The modules in the stack must have at least firmware from set A.06.34 and the above mentioned modules B.06.40 or above (must from the same firmware set).

- 1 Note the MAC address of the LAN interface (provided with G1369C LAN Interface Card or Main Board). This MAC address is on a label on the card or at the rear of the main board, e.g. *0030d3177321*.

On the Instant Pilot the MAC address can be found under **Details** in the LAN section.




Figure 62 LAN Setting on Instant Pilot

- 2 Set the Configuration Switch to DHCP either on the G1369C LAN Interface Card or the main board of above mentioned modules.

Table 27 G1369C LAN Interface Card (configuration switch on the card)

SW 4	SW 5	SW 6	SW 7	SW 8	Initialization Mode
ON	OFF	OFF	OFF	OFF	DHCP

Table 28 LC Modules inclusive 1120/1220 (configuration switch at rear of the instrument)

SW 6	SW 7	SW 8	Initialization Mode
ON	OFF	OFF	DHCP

- 3** Turn on the module that hosts the LAN interface.
- 4** Configure your Control Software (e.g. Agilent ChemStation, Lab Advisor, Firmware Update Tool) and use MAC address as host name, e.g. *0030d3177321*.

The LC system should become visible in the control software (see Note in section “[General Information \(DHCP\)](#)” on page 219).

Manual Configuration

Manual configuration only alters the set of parameters stored in the non-volatile memory of the module. It never affects the currently active parameters. Therefore, manual configuration can be done at any time. A power cycle is mandatory to make the stored parameters become the active parameters, given that the initialization mode selection switches are allowing it.




Figure 63 Manual Configuration (Principle)

With Telnet

Whenever a TCP/IP connection to the module is possible (TCP/IP parameters set by any method), the parameters may be altered by opening a Telnet session.

- 1 Open the system (DOS) prompt window by clicking on Windows **START** button and select **"Run..."**. Type "cmd" and press OK.
- 2 Type the following at the system (DOS) prompt:
 - `c:\>telnet <IP address>` or
 - `c:\>telnet <host name>`




```
C:\WINDOWS\system32\cmd.exe
C:\>telnet 134.40.30.205
```

Figure 64 Telnet - Starting a session

where <IP address> may be the assigned address from a Bootp cycle, a configuration session with the Handheld Controller, or the default IP address (see ["Configuration Switches"](#) on page 214).


When the connection was established successfully, the module responds with the following:



```
Telnet 134.40.30.205
Agilent Technologies G4212A PR00100015
>
```

Figure 65 A connection to the module is made

- 3 Type `?` and press enter to see the available commands.



```
Telnet 134.40.30.205
Agilent Technologies G4212A PR00100015
>?
command syntax      description
-----
?                    display help info
/                    display current LAN settings
ip <x.x.x.x>          set IP Address
sn <x.x.x.x>          set Subnet Mask
gw <x.x.x.x>          set Default Gateway
exit                exit shell
>
```

Figure 66 Telnet Commands

Table 29 Telnet Commands

Value	Description
?	displays syntax and descriptions of commands
/	displays current LAN settings
ip <x.x.x.x>	sets new ip address
sm <x.x.x.x>	sets new subnet mask
gw <x.x.x.x>	sets new default gateway
exit	exits shell and saves all changes

4 To change a parameter follows the style:

- parameter value, for example:
ip 134.40.28.56

Then press [Enter], where parameter refers to the configuration parameter you are defining, and value refers to the definitions you are assigning to that parameter. Each parameter entry is followed by a carriage return.

5 Use the “/” and press Enter to list the current settings.

```
cn Telnet 134.40.30.205
>/
LAN Status Page
-----
MAC Address   : 0030D317521C
Init Mode     : Using Stored
-----
TCP/IP Properties
- active -
IP Address    : 134.40.30.205
Subnet Mask   : 255.255.240.0
Def. Gateway  : 134.40.24.1
-----
TCP/IP Status : Ready
-----
Controllers   : no connections
>_
```

Figure 67 Telnet - Current settings in "Using Stored" mode

information about the LAN interface
MAC address, initialization mode
Initialization mode is Using Stored
active TCP/IP settings

TCP/IP status - here ready
connected to PC with controller software (e.g. Agilent
ChemStation), here not connected

- 6 Change the IP address (in this example 192.168.254.12) and type “/” to list current settings.

```

C:\Telnet 134.40.30.205
>ip 192.168.254.12
>/
LAN Status Page
-----
MAC Address      : 0030D317521C
Init Mode       : Using Stored
TCP/IP Properties
- active -
IP Address      : 134.40.30.205
Subnet Mask     : 255.255.240.0
Def. Gateway    : 134.40.24.1
- stored -
IP Address      : 192.168.254.12
Subnet Mask     : 255.255.240.0
Def. Gateway    : 134.40.24.1
TCP/IP Status   : Ready
Controllers     : no connections
>_

```

change of IP setting to
Initialization mode is Using Stored

active TCP/IP settings

stored TCP/IP settings in non-volatile memory

connected to PC with controller software (e.g. Agilent
ChemStation), here not connected

Figure 68 Telnet - Change IP settings

- 7 When you have finished typing the configuration parameters, type **exit** and press **Enter** to exit with storing parameters.

```

C:\WINDOWS\system32\cmd.exe
Agilent Technologies G4212A PR00100015
>exit

Connection to host lost.
C:\>_

```

Figure 69 Closing the Telnet Session

NOTE


If the Initialization Mode Switch is changed now to “Using Stored” mode, the instrument will take the stored settings when the module is re-booted. In the example above it would be 192.168.254.12.

PC and Agilent ChemStation Setup


PC Setup for Local Configuration

This procedure describes the change of the TCP/IP settings on your PC to match the module's default parameters in a local configuration (see Table 26 on page 218).

- 1 Open the Local Area Connection Properties and select **Internet Protocol (TCP/IP)**. Then click on **Properties**.




- 2 You may enter here the fixed IP address of the module or use the **Alternative Configuration**.



13 LAN Configuration

PC and Agilent ChemStation Setup

- 3 We will use the direct LAN access via Cross-over LAN cable with the module's IP address.



- 4 Click on **OK** to save the configuration.


Agilent ChemStation Setup

This procedure describes the Agilent ChemStation B.04.02 setup for the 1290 Infinity system using the 1290 Infinity DAD (G4212A) as the interfacing module. The setup works in the same way for all other systems.

NOTE

The LAN must be connected to detector due to high data load on communication to Control Software.

- 1 Open the ChemStation Configuration Editor.




- 2 Select from the menu **Configure - Instruments**.
- 3 Select **Modular 3D LC System**.
- 4 Give the Instrument a name.

13 LAN Configuration

PC and Agilent ChemStation Setup


5 Click on **OK**.



The **Select Instrument** dialog box is shown. It contains the following fields and options:

- Instrument Type:** A list box with the following items: LC 3D Data Analysis only, LC Data Analysis only, **Modular 3D LC System** (selected), Modular 3D LC System (Classic), Modular LC System, and Modular LC System (Classic).
- Instrument Name:** A text box containing the text "G4212A".
- Initially Start Instrument Session?:** Two radio buttons, **Yes** and **No**, both of which are unselected.
- Initial Screen Window Size:** Three radio buttons: **Normal** (selected), **Icon**, and **Full screen**.
- At the bottom are three buttons: **OK**, **Cancel**, and **Help**.

6 Select **LC System Access — Access Point** and click on **Add**.




The **G4212A - Device Configuration** dialog box is shown. It contains the following fields and options:

- Modular 3D LC System** (Title bar)
- Modules:** A list box with the following items: **LC System Access**, **Access Point** (selected), Dual Channel Iface, 35900E, CTCAutoSampler, and CTC PAL AutoSampler.
- GPIB Address:** A text box with a value of "1".
- GPIB:** A radio button that is selected.
- LAN:** A radio button that is unselected.
- Add:** A button next to the GPIB Address field.
- Selected Modules:** A list box with the following items: **LC System Access** and **Access Point**.
- Delete:** A button next to the Selected Modules list box.
- At the bottom are three buttons: **OK**, **Cancel**, and **Help**.

7 Click on **OK**.

The Configuration Editor shows now the new instrument.


- 8 If required, change under **Configure – Path** the folder locations.
- 9 Save the current configuration via **File – Save**.



- 10 Exit the Configuration Editor.
- 11 Start the Agilent ChemStation.

During first startup or when the system configuration has changed, a notification shows up.


- 12 The left column shows the modules that could be configured. You may select the module manually from the list. We use the Auto Configuration mode. Click on **Yes**.



13 LAN Configuration

PC and Agilent ChemStation Setup


- 13** Enter the IP address or the Hostname of the module with the LAN-access.



- 14** Click on **OK**.

The selected module is shown now in the right window (with serial number). In addition all other modules connected via CAN to the detector are shown as well.

- 15** Click on **OK** to continue the ChemStation loading.



- 16 You may see the details of the module by **selecting the module** and clicking on **Configure**.



Under **Connection Settings** you may change the IP/Hostname of the module (may require a re-start of the ChemStation).

After successful load of the ChemStation, you should see the module(s) as active item in the graphical user interface (GUI).

13 LAN Configuration
PC and Agilent ChemStation Setup





Figure 70 Screen After Successful Load of ChemStation



14 Appendix

General Safety Information	236
General Safety Information	236
Safety Standards	236
General	236
Before Applying Power	237
Ground the Instrument	237
Do Not Operate in an Explosive Atmosphere	238
Do Not Remove the Instrument Cover	238
Do Not Modify the Instrument	238
In Case of Damage	238
Solvents	239
Safety Symbols	240
Refrigerant	242
Waste Electrical and Electronic Equipment Directive	244
Radio Interference	245
Sound Emission	245
Solvent Information	246
Agilent Technologies on Internet	247

This chapter provides addition information on safety, legal and web.



General Safety Information

General Safety Information

The following general safety precautions must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the instrument. Agilent Technologies assumes no liability for the customer's failure to comply with these requirements.

WARNING

Ensure the proper usage of the equipment.

The protection provided by the equipment may be impaired.

→ The operator of this instrument is advised to use the equipment in a manner as specified in this manual.

Safety Standards

This is a Safety Class I instrument (provided with terminal for protective earthing) and has been manufactured and tested according to international safety standards.

General

Do not use this product in any manner not specified by the manufacturer. The protective features of this product may be impaired if it is used in a manner not specified in the operation instructions.

Before Applying Power

WARNING

Wrong voltage range, frequency or cabling

Personal injury or damage to the instrument

- Verify that the voltage range and frequency of your power distribution matches to the power specification of the individual instrument.
 - Never use cables other than the ones supplied by Agilent Technologies to ensure proper functionality and compliance with safety or EMC regulations.
 - Make all connections to the unit before applying power.
-

NOTE

Note the instrument's external markings described under “[Safety Symbols](#)” on page 240.

Ground the Instrument

WARNING

Missing electrical ground

Electrical shock

- If your product is provided with a grounding type power plug, the instrument chassis and cover must be connected to an electrical ground to minimize shock hazard.
 - The ground pin must be firmly connected to an electrical ground (safety ground) terminal at the power outlet. Any interruption of the protective (grounding) conductor or disconnection of the protective earth terminal will cause a potential shock hazard that could result in personal injury.
-

Do Not Operate in an Explosive Atmosphere

WARNING

Presence of flammable gases or fumes

Explosion hazard

- Do not operate the instrument in the presence of flammable gases or fumes.
-

Do Not Remove the Instrument Cover

WARNING

Instrument covers removed

Electrical shock

- Do Not Remove the Instrument Cover
 - Only Agilent authorized personnel are allowed to remove instrument covers. Always disconnect the power cables and any external circuits before removing the instrument cover.
-

Do Not Modify the Instrument

Do not install substitute parts or perform any unauthorized modification to the product. Return the product to an Agilent Sales and Service Office for service and repair to ensure that safety features are maintained.

In Case of Damage

WARNING

Damage to the module

Personal injury (for example electrical shock, intoxication)

- Instruments that appear damaged or defective should be made inoperative and secured against unintended operation until they can be repaired by qualified service personnel.
-

Solvents

WARNING

Toxic, flammable and hazardous solvents, samples and reagents

The handling of solvents, samples and reagents can hold health and safety risks.

- When working with these substances observe appropriate safety procedures (for example by wearing goggles, safety gloves and protective clothing) as described in the material handling and safety data sheet supplied by the vendor, and follow good laboratory practice.
 - The volume of substances should be reduced to the minimum required for the analysis.
 - Do not operate the instrument in an explosive atmosphere.
 - Never exceed the maximal permissible volume of solvents (6 L) in the solvent cabinet.
 - Do not use bottles that exceed the maximum permissible volume as specified in the usage guideline for the Agilent 1200 Infinity Series Solvent Cabinets.
 - Arrange the bottles as specified in the usage guideline for the solvent cabinet.
 - A printed copy of the guideline has been shipped with the solvent cabinet, electronic copies are available on the Internet.
 - Ground the waste container.
 - The residual free volume in the appropriate waste container must be large enough to collect the waste liquid.
 - Check the filling level of the waste container regularly.
 - To achieve maximal safety, check the correct installation regularly.
 - Do not use solvents with an auto-ignition temperature below 200 °C (392 °F).
-

Safety Symbols

Table 30 Symbols

	The apparatus is marked with this symbol when the user should refer to the instruction manual in order to protect risk of harm to the operator and to protect the apparatus against damage.
	Indicates dangerous voltages.
	Indicates a protected ground terminal.
	The apparatus is marked with this symbol when hot surfaces are available and the user should not touch it when heated up.
	Cooling unit is designed as vapor-compression refrigeration system. Contains fluorinated greenhouse gas (refrigerant) according to the Kyoto protocol. For specifications of refrigerant, charge capacity, carbon dioxide equivalent (CDE), and global warming potential (GWP) see instrument label.
	Confirms that a manufactured product complies with all applicable European Community directives. The European Declaration of Conformity is available at: http://regulations.corporate.agilent.com/DoC/search.htm
	Manufacturing date.
	Power symbol indicates On/Off. The apparatus is not completely disconnected from the mains supply when the power switch is in the Off position

Table 30 Symbols

	Pacemaker Magnets could affect the functioning of pacemakers and implanted heart defibrillators. A pacemaker could switch into test mode and cause illness. A heart defibrillator may stop working. If you wear these devices keep at least 55 mm distance to magnets. Warn others who wear these devices from getting too close to magnets.
	Magnetic field Magnets produce a far-reaching, strong magnetic field. They could damage TVs and laptops, computer hard drives, credit and ATM cards, data storage media, mechanical watches, hearing aids and speakers. Keep magnets at least 25 mm away from devices and objects that could be damaged by strong magnetic fields.
	Indicates a pinching or crushing hazard
	Indicates a piercing or cutting hazard.

WARNING

A WARNING

alerts you to situations that could cause physical injury or death.

- Do not proceed beyond a warning until you have fully understood and met the indicated conditions.

CAUTION

A CAUTION

alerts you to situations that could cause loss of data, or damage of equipment.

- Do not proceed beyond a caution until you have fully understood and met the indicated conditions.

Refrigerant

The refrigerant HFKW-134a is used only in the Agilent Infinity II Sample Cooler.

Table 31 Physical properties of refrigerant HFKW-134a

Molecular weight	102
Critical temperature	101.1 °C
Critical pressure	40.6 bar
Boiling point	-26.5 °C

WARNING

Refrigerant



Refrigerant HFKW-134a is known as a safe refrigerant, however accidents can occur if it is handled incorrectly. For this reason, the following instructions must be observed:

- Avoid contact with liquid refrigerant HFKW-134a. At atmospheric pressure HFKW-134a evaporates at approximately -26 °C and causes frost bite.
- After skin contact, rinse the affected area with water.
- After eye contact, rinse the eye(s) with plenty of water for at least 15 minutes and consult a doctor.
- HFKW-134a must not be allowed to escape in enclosed areas. Although HFKW-134a is not toxic, there is a danger of suffocation as gaseous refrigerant is heavier than air.
- Please observe the following first aid instructions. After inhalation, move the affected person to fresh air, keep him warm and allow him to rest. If necessary, he should be supplied with oxygen. If he has stopped breathing or is breathing erratically, he should be given artificial respiration. In the case of cardiac arrest, carry out heart massage. Send for a doctor immediately.
- Moreover, it must be noted that HFKW-134a must always be extracted from the system and collected. It must never be discharged into the atmosphere on environmental grounds (greenhouse effect).

CAUTION

General hazards and improper disposal

Improper disposal of the media and components used pollutes the environment.

- The breakdown of the sample cooler unit must be carried out by specialist refrigeration company.
 - All media must be disposed of in accordance with national and local regulations.
 - Please contact your local Agilent Service Center in regard to safe environmental disposal of the appliance.
-

Waste Electrical and Electronic Equipment Directive

Abstract

The Waste Electrical and Electronic Equipment (WEEE) Directive (2002/96/EC), adopted by EU Commission on 13 February 2003, is introducing producer responsibility on all electric and electronic appliances starting with 13 August 2005.

NOTE

This product complies with the WEEE Directive (2002/96/EC) marking requirements. The affixed label indicates that you must not discard this electrical/electronic product in domestic household waste.

Product Category:

With reference to the equipment types in the WEEE Directive Annex I, this product is classed as a Monitoring and Control Instrumentation product.



NOTE

Do not dispose of in domestic household waste

To return unwanted products, contact your local Agilent office, or see www.agilent.com for more information.

Radio Interference

Cables supplied by Agilent Technologies are screened to provide optimized protection against radio interference. All cables are in compliance with safety or EMC regulations.

Test and Measurement

If test and measurement equipment is operated with unscreened cables, or used for measurements on open set-ups, the user has to assure that under operating conditions the radio interference limits are still met within the premises.

Sound Emission

Manufacturer's Declaration

This statement is provided to comply with the requirements of the German Sound Emission Directive of 18 January 1991.

This product has a sound pressure emission (at the operator position) < 70 dB.

- Sound Pressure $L_p < 70$ dB (A)
- At Operator Position
- Normal Operation
- According to ISO 7779:1988/EN 27779/1991 (Type Test)

Solvent Information

Flow Cell


To protect optimal functionality of your flow-cell:

- The recommended pH range of the cell is 1.0 - 12.5 (solvent dependent).
- If the flow cell is transported while temperatures are below 5 degree C, it must be assured that the cell is filled with alcohol.
- Aqueous solvents in the flow cell can build up algae. Therefore do not leave aqueous solvents sitting in the flow cell. Add a small % of organic solvents (e.g. acetonitrile or methanol ~5%).

Use of Solvents

Observe the following recommendations on the use of solvents.

- Brown glass ware can avoid growth of algae.
- Avoid the use of the following steel-corrosive solvents:
 - Solutions of alkali halides and their respective acids (for example, lithium iodide, potassium chloride, and so on),
 - High concentrations of inorganic acids like sulfuric acid and nitric acid, especially at higher temperatures (if your chromatography method allows, replace by phosphoric acid or phosphate buffer which are less corrosive against stainless steel),
 - Halogenated solvents or mixtures which form radicals and/or acids, for example:



This reaction, in which stainless steel probably acts as a catalyst, occurs quickly with dried chloroform if the drying process removes the stabilizing alcohol,

- Chromatographic grade ethers, which can contain peroxides (for example, THF, dioxane, di-isopropyl ether) such ethers should be filtered through dry aluminium oxide which adsorbs the peroxides,
- Solvents containing strong complexing agents (e.g. EDTA),
- Mixtures of carbon tetrachloride with 2-propanol or THF.

Agilent Technologies on Internet

For the latest information on products and services visit our worldwide web site on the Internet at:

<http://www.agilent.com>

Index

A

accessory kit parts 176
 Agilent Lab Advisor software 99
 Agilent Lab Advisor 99
 Agilent
 ChemStation Setup 226
 on internet 247
 algae 246, 246
 ambient non-operating temperature 27
 ambient operating temperature 27
 analog signal 198
 array
 diode 17
 astm drift 133
 ASTM
 environmental conditions 26

B

bandwidth 90
 Beer-Lambert (law) 72
 bench space 26
 Bootp
 & Store 216
 using default 217
 using stored 217

C

cable
 CAN 186
 LAN 186
 overview 178
 RS-232 187
 cables
 analog 180

remote 182
 CAN
 cable 186
 cartridge
 cleaning 163
 max-light 163
 cautions and warnings 152
 cell test 127
 ChemStation
 setup 229
 chromatogram 55
 clean
 Max-light cartridge cell 163
 cleaning
 flow cell 163
 compatibility 170
 compensation sensor open 107
 compensation sensor short 108
 condensation 26
 configuration switch 214
 conversion 170
 correction factors for flow cells 73

D

D/A converter 144
 DAC 144
 dark current 147
 data collection rate 91
 delay volume 87
 detection
 compound classes 85
 detector set up
 Open Lab ChemStation 39
 detector warm up 94

detector
 achieving higher sensitivity 88
 preparation 56
 DHCP
 general information 219
 setup 220
 dimensions 27
 doors
 install 156, 156
 remove 156, 156

E

electrical connections
 descriptions of 193
 electronic waste 244
 EMF
 early maintenance feedback 208
 emulation 170
 environment 26
 error messages
 compensation sensor open 107
 compensation sensor short 108
 diode current leakage 111
 fan failed 108
 heater failed 115
 heater power at limit 116
 ignition without cover 109, 109
 illegal temperature value from sensor
 at air inlet 115
 illegal value from sensor on main
 board 114
 leak sensor open 107
 leak sensor short 106
 leak 109
 lost CAN partner 106

- remote timeout 105
- shutdown 104
- timeout 104
- uv heater current 113
- uv ignition failed 112
- uv lamp current 111
- uv lamp voltage 112
- wavelength calibration failed 113
- wavelength recalibration lost 114
- extra-column volume
 - description 87

F

- fan failed 108
- firmware
 - description 190
 - initial 170
 - main system 190
 - resident system 190
 - update tool 191
 - updates 191, 169
 - upgrade/downgrade 169
- flow cell
 - correction factors 73
 - Max-Light cartridge flow cell 88
 - Max-Light high sensitivity cell 88
 - solvent information 246
- frequency range 27

G

- general error messages 104
- grating 17

H

- humidity 27

I

- inline pressure relief valve kit 70
- installation

- bench space 26
- environment 26
- power considerations 24

- install
 - doors 156
- instrument layout 207
- intensity test 124
- interfaces
 - Infinity II 196
- internet 247

L

- lamp
 - uv 14
- LAN configuration
 - ChemStation 229
 - PC setup 226
- LAN
 - Bootp & Store 216
 - Bootp 215
 - cable 186
 - configuration switch 214
 - manual configuration with telnet 223
 - manual configuration 222
 - PC and Agilent ChemStation Setup 226
 - TCP/IP parameter configuration 213
 - using default 217
 - using stored 217
 - what you have to do first 212, 212
- leak parts
 - replace 167
- leak sensor open 107
- leak sensor short 106
- leak 109
- line frequency 27
- line voltage 27
- linearity
 - specifications 32

- local configuration 226
- lost CAN partner 106

M

- MAC
 - address 212
- maintenance
 - definition of 152
 - drying the leak sensor 166
 - feedback 208
 - lamp and flow cell RFID tag 171
 - overview 154
 - replacing firmware 169
 - replacing the deuterium lamp 157
 - replacing the flow cell 160
 - serial number and firmware revision 171
- manual configuration
 - of LAN 222
- Max-light cartridge cell
 - clean 163
 - install 160
 - remove 160
 - replace 160
 - store 165
- message
 - diode current leakage 111
 - heater failed 115
 - heater power at limit 116
 - ignition without cover 109, 109
 - illegal temperature value from sensor at air inlet 115
 - illegal value from sensor on main board 114
 - remote timeout 105
 - uv heater current 113
 - uv ignition failed 112
 - uv lamp current 111
 - uv lamp voltage 112
 - wavelength calibration failed 113

Index

module firmware
 replace 169

N

negative absorbance 82
Noise and Linearity
 specifications 32
noise 133
non-operating altitude 27
non-operating temperature 27

O

operating Altitude 27
operating principle 21
operating temperature 27
optical 13
 system 13
optimization
 detector performance 66
 detector sensitivity 88
 for sensitivity, selectivity, linearity,
 dispersion 72
 how to get the best performance 66
 margins for negative absorbance 82
 of selectivity 83
 peak width 73
 performance 65
 sample and reference
 wavelength 77
 slit width 79
 spectra acquisition 81
 wavelength and bandwidth 88
optimizing
 1260 System 87
 flow cell 69
overpressure relieve valve kit 70
overview
 cable 178

P

parts identification
 accessory kit 176
PC and Agilent ChemStation Setup 226
PC setup
 local configuration 226
peak width (response time) 73
peak width 91
performance specifications 28
performance
 optimization 66
 specifications 30, 28
photometric accuracy 73
physical specifications 27
power considerations 24
power consumption 27
power cords 25
product description 11, 12

Q

quick noise test 130

R

radio interference 245
remote (ERI) 198
remote
 cables 182
remove
 doors 156
repairs
 cautions and warnings 152
 replacing firmware 169
replace
 leak parts 167
 max-light cartridge cell 160
response time (peak width) 73
response time versus time constant 32
response time 91

RS-232C
 cable 187
running the sample 58

S

safety class I 236
safety
 general information 236
 standards 27
 symbols 240
sample and reference wavelength 77
selectivity optimization 83
self test 122
serial number
 information 194, 195
set up
 detector 39
setting up an analysis 52
settings
 method parameters 43
setup
 ChemStation 229
shutdown 104
signal wavelength 90
site requirements
 power cords 25
slit test 136
slit width 79
slit
 fixed 16
 programmable 15
solvents 246
sound emission 245
special interfaces 199
special settings
 boot main 205
 boot-resident 205
specification
 physical 27

- specifications 32
 - noise and linearity 32
 - performance 30, 28
 - spectra
 - acquisition 81
 - store
 - Max-light cartridge cell 165
 - suppression
 - quantifying 83
 - system 13
 - optical 13
- ## T
- TCP/IP parameter configuration 213
 - telnet
 - configuration 223
 - temperature sensor 109
 - test cell
 - usage 120
 - test function
 - astm drift 133
 - cell test 127
 - conditions 121
 - D/A converter 144
 - DAC 144
 - dark current 147
 - failing 121
 - intensity test 124
 - introduction 118
 - max-light cartridge 120
 - noise 133
 - quick noise test 130
 - self test 122
 - slit 136
 - test cell cartridge 120
 - wavelength calibration 141
 - wavelength verification 139
 - tests
 - available tests vs interface 98
 - time constant versus response time 32
 - timeout 104
 - troubleshooting
 - available tests vs interface 98
 - error messages 103
- ## U
- USB 197
 - Using
 - detector 56
 - priming and purging the system 52
 - requirements and conditions 54
 - running the sample 58
 - setting up an analysis 52
 - typical chromatogram 55
- ## V
- voltage range 27
- ## W
- warm up of detector 94
 - warnings and cautions 152
 - waste
 - electrical and electronic equipment 244
 - wavelength and bandwidth
 - optimization 88
 - wavelength calibration 141
 - wavelength recalibration lost 114
 - wavelength verification
 - test 139
 - WEEE directive 244
 - weight 27

In This Book

This manual contains technical reference information about the Agilent 1290 Infinity II Diode Array Detector FS (G7117A) and the Agilent 1290 Infinity II Diode Array Detector (G7117B).

The manual describes the following:

- introduction and specifications,
- using and optimizing,
- troubleshooting and diagnose,
- maintenance,
- parts identification,
- hardware information,
- safety and related information.

© Agilent Technologies 2014

Printed in Germany
12/2014



G7117-90000
Rev. B